

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

|  |  |  |
|--|--|--|
|  <p><b>UKAS</b><br/>CALIBRATION</p> <p><b>0295</b></p> <p>Accredited to<br/><b>ISO/IEC 17025:2017</b></p> | <p><b>Haven Automation Limited</b></p> <p><b>Issue No: 054    Issue date: 10 October 2023</b></p>  |  |
|  | <p><b>Measurement House</b></p> <p><b>Kingsway</b></p> <p><b>Fforestfach</b></p> <p><b>Swansea</b></p> <p><b>Wales</b></p> <p><b>SA5 4EX</b></p> | <p><b>Contact: Mr David Gray</b></p> <p><b>Tel: +44 (0)1792 588722</b></p> <p><b>Fax: +44 (0)1792 582624</b></p> <p><b>E-Mail: mail@haven.co.uk</b></p> <p><b>Website: www.haven.co.uk</b></p> |
| <p><b>Calibration performed at the above address only</b></p>  |  |  |

### Calibration and Measurement Capability (CMC)

| Measured Quantity<br>Instrument or Gauge | Range  | Expanded<br>Measurement<br>Uncertainty<br>( $k = 2$ )  | Remarks  |
|--|--|--|--|
| <b>ELECTRICAL</b>                        |  |  |  |
| DC Resistance<br>Measurement             | 0 $\Omega$ to 20 $\Omega$<br>20 $\Omega$ to 200 $\Omega$<br>200 $\Omega$ to 2 k $\Omega$<br>2 k $\Omega$ to 20 k $\Omega$<br>20 k $\Omega$ to 200 k $\Omega$<br>200 k $\Omega$ to 2 M $\Omega$<br>2 M $\Omega$ to 20 M $\Omega$<br>20 M $\Omega$ to 200 M $\Omega$<br>200 M $\Omega$ to 1 G $\Omega$ | 25 $\mu\Omega/\Omega$ + 70 $\mu\Omega$<br>15 $\mu\Omega/\Omega$<br>10 $\mu\Omega/\Omega$<br>12 $\mu\Omega/\Omega$<br>20 $\mu\Omega/\Omega$<br>25 $\mu\Omega/\Omega$<br>0.010 %<br>0.020 % + 15 k $\Omega$<br>0.060 % + 0.60 M $\Omega$ | Generation of these quantities<br>with the same or similar CMCs<br>may be undertaken over the<br>same ranges by the use of a<br>transfer method. |
| Generation                               | 0 $\Omega$ to 40 $\Omega$<br>40 $\Omega$ to 400 $\Omega$<br>400 $\Omega$ to 4 k $\Omega$<br>4 k $\Omega$ to 40 k $\Omega$<br>40 k $\Omega$ to 400 k $\Omega$<br>400 k $\Omega$ to 4 M $\Omega$<br>4 M $\Omega$ to 40 M $\Omega$<br>40 M $\Omega$ to 400 M $\Omega$                                   | 0.018 % + 100 $\mu\Omega$<br>0.0085 %<br>0.0055 %<br>0.0075 %<br>0.0095 %<br>0.012 %<br>0.030 %<br>0.024 %   | Using multi-function calibrator.   |
| DC Voltage<br>Measurement                | 0 mV to 200 mV<br>200 mV to 2 V<br>2 V to 20 V<br>20 V to 200 V<br>200 V to 1 kV   | 1.5 $\mu\text{V}$<br>10 $\mu\text{V}/\text{V}$<br>8.0 $\mu\text{V}/\text{V}$<br>10 $\mu\text{V}/\text{V}$<br>15 $\mu\text{V}/\text{V}$   | Generation of these quantities<br>with the same or similar CMCs<br>may be undertaken over the<br>same ranges by the use of a<br>transfer method. |
| Generation                               | 1 kV to 5 kV<br>0 mV to 320 mV<br>320 mV to 3.2 V<br>3.2 V to 32 V<br>32 V to 320 V<br>320 V to 1050 V   | 1.2 %<br>0.0045 % + 2.5 $\mu\text{V}$<br>0.0060 %<br>0.0070 %<br>0.0090 %<br>0.0060 %  | Using multi-function calibrator.   |



0295  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Haven Automation Limited**  
**Issue No: 054 Issue date: 10 Octobr 2023**

Calibration performed at main address only

| Measured Quantity<br>Instrument or Gauge | Range   | Expanded<br>Measurement<br>Uncertainty<br>( $k = 2$ )                   | Remarks  |
|--|---|---|--|
| <b>ELECTRICAL (continued)</b>            |   |   |  |
| <b>DC Current</b>                        |   |   |  |
| Measurement                              | 0 $\mu$ A to 200 $\mu$ A<br>0.2 mA to 20 mA<br>20 mA to 200 mA<br>200 mA to 2 A   | 0.0065 % + 10 nA<br>0.0055 % + 10 nA<br>0.0080 %<br>0.015 %             | Generation of these quantities with the same or similar CMCs may be undertaken over the same ranges by the use of a transfer method. |
| Generation                               | 0 $\mu$ A to 320 $\mu$ A<br>320 $\mu$ A to 3.2 mA<br>3.2 mA to 32 mA<br>32 mA to 320 mA<br>320 mA to 3 A<br>3 A to 10 A | 0.012 % + 6.0 nA<br>0.021 %<br>0.026 %<br>0.040 %<br>0.040 %<br>0.042 % | Using multi-function calibrator.   |
| <b>AC Voltage</b>                        |   |   |  |
| Generation                               | 40 Hz to 30 kHz<br>32 mV to 320 mV<br>320 mV to 320 V   | 0.040 %<br>0.050 %  | Using multi-function calibrator.   |
| Measurement                              | 40 Hz to 10 kHz<br>320 V to 750 V<br>750 V to 1050 V  | 0.040 %<br>0.050 %  |  |
| Measurement                              | 40 Hz to 30 kHz<br>200 mV to 2 V<br>2 V to 20 V<br>20 V to 200 V  | 0.030 %<br>0.030 %<br>0.028 %   | Generation of these quantities with the same or similar CMCs may be undertaken over the same ranges by the use of a transfer method. |
| Measurement                              | 40 Hz to 10 kHz<br>20 mV to 200 mV<br>200 V to 1000 V   | 59 $\mu$ V<br>0.040 %   |  |
| <b>AC Current</b>                        |   |   |  |
| Generation                               | 50 Hz<br>1 kV to 5 kV   | 1.8 %   |  |
| Generation                               | 32 $\mu$ A to 320 mA<br>10 Hz to 110 Hz<br>110 Hz to 3 kHz  | 0.045 %<br>0.070 %  | Using multi-function calibrator.   |
| Generation                               | 320 mA to 3 A<br>40 Hz to 110 Hz<br>110 Hz to 3 kHz   | 0.060 %<br>0.080 %  |  |
| Generation                               | 3 A to 10 A<br>40 Hz to 110 Hz<br>110 Hz to 3 kHz   | 0.080 %<br>0.080 %  |  |



0295  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Haven Automation Limited**  
**Issue No: 054 Issue date: 10 Octobr 2023**

Calibration performed at main address only

| Measured Quantity<br>Instrument or Gauge                           | Range   | Expanded<br>Measurement<br>Uncertainty<br>( <i>k</i> = 2)                    | Remarks  |
|--|---|--|--|
| ELECTRICAL (continued)   |   |  |  |
| AC Current (continued)   |   |  |  |
| Measurement  | 40 Hz to 1 kHz<br>100 µA to 200 µA<br>0.2 mA to 2 mA<br>2 mA to 20 mA<br>20 mA to 200 mA  | 0.035 µA<br>0.030 % + 0.25 µA<br>0.030 % + 2.5 µA<br>0.030 % + 25 µA         | Generation of these quantities with the same or similar CMCs may be undertaken over the same ranges by the use of a transfer method. |
|  | 55 Hz to 300 Hz<br>200 mA to 500 mA<br>500 mA to 2 A  | 0.25 %<br>0.10 %   |  |
| AC Resistance  | At 50 Hz<br>0.05 Ω and 0.1 Ω<br>0.2 Ω, 0.5 Ω, 1 Ω, 5 Ω, 10 Ω,<br>50 Ω, 100 Ω, 500 Ω and 1 kΩ  | 0.40 %<br>0.30 %   | For the calibration of the earth bond function on Portable Appliance Testers.  |
| Capacitance  |   |  |  |
| Generation   | 1 nF to 4 nF<br>4 nF to 40 nF<br>40 nF to 400 nF<br>400 nF to 4 µF<br>4 µF to 40 µF<br>40 µF to 400 µF<br>400 µF to 4 mF<br>4 mF to 30 mF | 0.30 %<br>0.20 %<br>0.20 %<br>0.20 %<br>0.20 %<br>0.20 %<br>0.35 %<br>0.35 % |  |
| Frequency  |   |  |  |
| Generation   | 0.5 Hz to 200 kHz   | 0.0012 % + 0.010 Hz  |  |
| Temperature indicators,<br>calibration by electrical<br>simulation |   |  |  |
| Base metal thermocouple  | -50 °C to +1320 °C  | 0.30 °C  | Including cold junction compensation   |
| Noble metal thermocouple   | -50 °C to +1800 °C  | 0.70 °C  | Including cold junction compensation   |
| Resistance thermometer<br>(Pt 100)                                 | -200 °C to 0 °C<br>0 °C to 250 °C<br>250 °C to 800 °C   | 0.0020 °C to 0.012 °C<br>0.012 °C to 0.025 °C<br>0.025 °C to 0.030 °C        |  |
| Cold junction compensation   | 21 °C to 25 °C  | 0.15 °C  | Lab ambient temperature  |



0295  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Haven Automation Limited**  
**Issue No: 054 Issue date: 10 Octobr 2023**

Calibration performed at main address only

| Measured Quantity<br>Instrument or Gauge                           | Range   | Expanded<br>Measurement<br>Uncertainty<br>( $k = 2$ )                 | Remarks   |
|--|---|---|---|
| <b>ELECTRICAL (continued)</b>                                      |   |   |   |
| Temperature simulators,<br>calibration by electrical<br>simulation |   |   |   |
| Base metal thermocouple  | -50 °C to +1320 °C  | 0.35 °C   | Including cold junction<br>compensation   |
| Noble metal thermocouple   | -50 °C to +1800 °C  | 0.90 °C   | Including cold junction<br>compensation   |
| Resistance thermometer<br>(Pt 100)                                 | -200 °C to 0 °C<br>0 °C to 250 °C<br>250 °C to 800 °C   | 0.0020 °C to 0.012 °C<br>0.012 °C to 0.025 °C<br>0.025 °C to 0.030 °C |   |
| Cold junction compensation   | 21 °C to 25 °C  | 0.15 °C   | Lab ambient temperature   |
| <b>PRESSURE</b>  |   |   |   |
| <u>Hydraulic pressure (gauge)</u>                                  |   |   |   |
| Calibration of pressure indicating<br>instruments and gauges       | 552 kPa to 4.8 MPa<br>5.1 MPa to 110 MPa  | 0.0090 % + 0.15 kPa<br>0.010 %  | 1 Calibrations may be<br>undertaken expressed in other<br>units of pressure as required.<br><br>2 Calibration of pressure<br>measuring devices with an<br>electrical output may be<br>undertaken. |
| <u>Gas pressure (gauge)</u>  |   |   |   |
| Calibration of pressure indicating<br>instruments and gauges       | -90 kPa to -2.5 kPa<br>1.5 kPa to 2.5 kPa<br>2.5 kPa to 100 kPa<br>100 kPa to 690 kPa<br>690 kPa to 3.5 MPa | 0.011 %<br>0.032 %<br>0.011 %<br>0.0060 %<br>0.0080 %                 |   |
| <b>TEMPERATURE</b>   |   |   |   |
| Resistance Thermometers  | -50 °C to 0 °C<br><br>0 °C<br>0 °C to 230 °C<br>230 °C to 420 °C<br>420 °C to 650 °C                        | 0.070 °C<br><br>0.060 °C<br>0.10 °C<br>0.10 °C<br>0.10 °C             | Calibration performed within<br>Liquid Baths  |
| Thermocouples  |   |   |   |
| Base metal<br>(Type J, K, T & N)                                   | -50 °C to +650 °C   | 0.42 °C   |   |



0295  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Haven Automation Limited**  
**Issue No: 054 Issue date: 10 Octobr 2023**

Calibration performed at main address only

| Measured Quantity<br>Instrument or Gauge | Range                                 | Expanded<br>Measurement<br>Uncertainty<br>( $k = 2$ ) | Remarks   |
|--|---------------------------------------|---|---|
| TEMPERATURE (continued)                  |                                       |   |   |
| Electronic thermometers with sensors     |                                       | As for sensor type                                    |   |
| Temperature loggers with integral probes | -50 °C to 0 °C<br>0 °C to 150 °C      | 0.070 °C<br>0.10 °C                                   |   |
| Metal Block calibrators                  | -50 °C to +300 °C<br>300 °C to 650 °C | 0.20 °C<br>0.35 °C                                    | Method consistent with Euramet CG13   |
| Liquid Baths                             | -40 °C to +250 °C                     | 0.20 °C   |   |
| Furnaces and ovens                       | 50 °C to 600 °C                       | 3.0 °C  | Single or Multipoint monitoring probes.<br>Time dependent temperature profiling, also referred to as spatial temperature surveying or mapping |
| END                                      |                                       |   |   |



0295  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Haven Automation Limited**  
**Issue No: 054 Issue date: 10 Octobr 2023**

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$