

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

|  |  |   |
|--|--|---|
|  <p><b>UKAS</b><br/>CALIBRATION</p> <p><b>0461</b></p> <p>Accredited to<br/><b>ISO/IEC 17025:2017</b></p> | <p><b>TMS Europe Ltd</b></p> <p>Issue No: 060      Issue date: 30 May 2024</p> |   |
|  | <p>Unit 10<br/>Stretfield Mill<br/>Bradwell<br/>Hope Valley<br/>S33 9JT</p>    | <p>Contact: Mr S Nuttall<br/>Tel: +44 (0)1433 620535<br/>Fax: +44 (0)1433 621492<br/>E-Mail: sales@tmseurope.co.uk<br/>Website: www.tmseurope.co.uk</p> |

Calibration performed by the Organisations at the locations specified below

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

| Location details  | Activity   | Location code |
|---|--|---------------|
| <p><b>Address</b><br/>Unit 10<br/>Stretfield Mill<br/>Bradwell<br/>Hope Valley<br/>S33 9JT</p> <p><b>Local contact</b><br/>Mr B Hanwell<br/>Tel: +44 (0)1433 620535<br/>Fax: +44 (0)1433 621492<br/>Email: sales@tmseurope.co.uk<br/>Website: www.tmseurope.co.uk</p> | <p><a href="#">Electrical</a><br/><a href="#">Time</a><br/><a href="#">Humidity</a><br/><a href="#">Pressure</a><br/><a href="#">Temperature</a></p> | P             |

#### Site activities performed away from the locations listed above:

| Location details   | Activity   | Location code |
|--|--|---------------|
| <p>The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer</p> <p>Contact as above</p> | <p><a href="#">Electrical</a><br/><a href="#">Time</a><br/><a href="#">Pressure</a><br/><a href="#">Temperature</a><br/><a href="#">Humidity</a></p> | S             |



0461  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TMS Europe Ltd**  
Issue No: 060 Issue date: 30 May 2024

Calibration performed by the Organisation at the locations specified

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)

| Measured Quantity<br>Instrument or Gauge | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )  | Remarks  | Location<br>Code |
|--|---|---|--|------------------|
| <b>ELECTRICAL CALIBRATION</b>            |   |   |  |                  |
| <b>DC VOLTAGE</b>                        |   |   |  |                  |
| Measurement and Generation               | 0 mV to 100 mV<br>100 mV to 1 V<br>1 V to 10 V<br>10 V to 100 V<br>100 V to 1000 V  | 0.0020 % + 0.50 $\mu$ V<br>0.0020 % + 2.0 $\mu$ V<br>0.0010 % + 7.0 $\mu$ V<br>0.0015 % + 120 $\mu$ V<br>0.0015 % + 2.5 mV  | Known values of DC Voltage for application to measuring instruments or measurement of DC voltages supplied by generators, power supplies etc.                            | P                |
| <b>DC CURRENT</b>                        |   |   |  |                  |
| Measurement                              | 0 $\mu$ A to 10 $\mu$ A<br>10 $\mu$ A to 100 $\mu$ A<br>100 $\mu$ A to 1 mA<br>1 mA to 10 mA<br>10 mA to 100 mA   | 0.0045 % + 3.8 nA<br>0.0045 % + 11 nA<br>0.0045 % + 110 nA<br>0.0040 % + 1.0 $\mu$ A<br>0.0039 % + 10 $\mu$ A   | Voltage and resistance method  | P                |
| Generation                               | 0 mA to 100 $\mu$ A<br>100 $\mu$ A to 1 mA<br>1 mA to 10 mA<br>10 mA to 100 mA<br>100 mA to 1 A<br>1 A to 3 A   | 60 $\mu$ A/A + 0.010 $\mu$ A<br>60 $\mu$ A/A + 0.040 $\mu$ A<br>60 $\mu$ A/A + 0.40 $\mu$ A<br>120 $\mu$ A/A + 6.0 $\mu$ A<br>170 $\mu$ A/A + 0.12 mA<br>0.14 % + 0.90 mA                             | Using digital multimeter   |                  |
| Measurement and Generation               | 0 $\mu$ A to 200 $\mu$ A<br>200 $\mu$ A to 2 mA<br>2 mA to 20 mA<br>20 mA to 200 mA<br>200 mA to 2 A<br>2 A to 30 A   | 0.012 % + 37 nA<br>0.0095 % + 60 nA<br>0.0060 % + 0.45 $\mu$ A<br>0.0095 % + 5.0 $\mu$ A<br>0.018 % + 110 $\mu$ A<br>0.047 % + 4.6 mA   | Using multi-function calibrator.   | P                |
| <b>DC RESISTANCE</b>                     |   |   |  |                  |
| Measurement and Generation               | 0 $\Omega$ to 10 $\Omega$<br>10 $\Omega$ to 100 $\Omega$<br>100 $\Omega$ to 1 k $\Omega$<br>1 k $\Omega$ to 10 k $\Omega$<br>10 k $\Omega$ to 100 k $\Omega$<br>100 k $\Omega$ to 1 M $\Omega$<br>1 M $\Omega$ to 10 M $\Omega$ | 0.0040 % + 46 $\mu$ $\Omega$<br>0.0020 % + 1.3 m $\Omega$<br>0.0020 % + 1.5 m $\Omega$<br>0.0018 % + 12 m $\Omega$<br>0.0020 % + 120 m $\Omega$<br>0.0025 % + 2.5 $\Omega$<br>0.0056 % + 120 $\Omega$ | Known values of DC Resistance for application to measuring instruments or measurement of DC resistances supplied by resistors, resistance boxes and similar instruments. | P                |
| Generation                               | 0.1 $\Omega$<br>1 $\Omega$<br>10 $\Omega$<br>100 $\Omega$<br>1 k $\Omega$<br>10 k $\Omega$<br>100 k $\Omega$<br>1 M $\Omega$<br>10 M $\Omega$<br>100 M $\Omega$   | 6.0 m $\Omega$<br>6.0 m $\Omega$<br>7.0 m $\Omega$<br>12 m $\Omega$<br>95 m $\Omega$<br>0.93 $\Omega$<br>9.4 $\Omega$<br>170 $\Omega$<br>4.6 k $\Omega$<br>580 k $\Omega$                             | Known, fixed values of DC resistance for application to resistance measuring devices.  |                  |



0461  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TMS Europe Ltd**  
Issue No: 060 Issue date: 30 May 2024

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range  | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )   | Remarks                          | Location<br>Code |
|--|--|--|----------------------------------|------------------|
| AC VOLTAGE                               |  |  |                                  |                  |
| Measurement                              | 10 Hz to 20 kHz<br>100 $\mu$ V to 100 mV<br>100 mV to 1 V<br>1 V to 10 V<br>10 V to 100 V<br>100 V to 750 V      | 0.070 % + 60 $\mu$ V<br>0.070 % + 500 $\mu$ V<br>0.070 % + 5.5 mV<br>0.070 % + 45 mV<br>0.070 % + 360 mV   | Using digital multimeter.        | P                |
| Generation                               | 0 mV to 200 mV<br>10 Hz to 44 Hz<br>44 Hz to 1 kHz<br>1 kHz to 20 kHz<br>20 kHz to 100 kHz<br>100 kHz to 500 kHz | 0.24 % + 90 $\mu$ V<br>0.050 % + 50 $\mu$ V<br>0.11 % + 220 $\mu$ V<br>0.35 % + 2.6 mV<br>0.95 % + 2.6 mV  | Using multi-function calibrator. | P                |
|  | 200 mV to 2 V<br>10 Hz to 44 Hz<br>44 Hz to 1 kHz<br>1 kHz to 20 kHz<br>20 kHz to 100 kHz<br>100 kHz to 500 kHz  | 0.24 % + 750 $\mu$ V<br>0.050 % + 420 $\mu$ V<br>0.11 % + 580 $\mu$ V<br>0.30 % + 3.2 mV<br>0.52 % + 26 mV |                                  |                  |
|  | 2 V to 20 V<br>10 Hz to 44 Hz<br>44 Hz to 1 kHz<br>1 kHz to 20 kHz<br>20 kHz to 100 kHz                          | 0.24 % + 8.3 mV<br>0.042 % + 7.0 mV<br>0.084 % + 7.1 mV<br>0.26 % + 43 mV                                  |                                  |                  |
|  | 20 V to 200 V<br>30 Hz to 44 Hz<br>44 Hz to 1 kHz<br>1 kHz to 10 kHz   | 0.070 % + 120 mV<br>0.046 % + 120 mV<br>0.11 % + 310 mV  |                                  |                  |
|  | 200 V to 1000 V<br>40 Hz to 44 Hz<br>44 Hz to 1 kHz<br>1 kHz to 10 kHz   | 0.070 % + 420 mV<br>0.046 % + 500 mV<br>0.18 % + 860 mV  |                                  |                  |
| AC CURRENT                               |  |  |                                  |                  |
| Generation                               | 20 $\mu$ A to 200 $\mu$ A<br>10 Hz to 44 Hz<br>44 Hz to 1 kHz<br>1 kHz to 10 kHz                                 | 0.23 % + 0.40 $\mu$ A<br>0.082 % + 0.32 $\mu$ A<br>0.93 % + 0.70 $\mu$ A                                   | Using multi-function calibrator. | P                |
|  | 0.2 mA to 2 mA<br>10 Hz to 44 Hz<br>44 Hz to 1 kHz<br>1 kHz to 10 kHz  | 0.23 % + 5.0 $\mu$ A<br>0.070 % + 0.80 $\mu$ A<br>0.81 % + 7.0 $\mu$ A                                     |                                  |                  |



0461  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TMS Europe Ltd**  
Issue No: 060 Issue date: 30 May 2024

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge   | Range  | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )  | Remarks  | Location<br>Code |
|--|--|---|--|------------------|
| AC CURRENT (continued)<br>Generation (continued)   | 2 mA to 20 mA<br>10 Hz to 44 Hz<br>44 Hz to 1 kHz<br>1 kHz to 10 kHz<br><br>20 mA to 200 mA<br>10 Hz to 44 Hz<br>44 Hz to 1 kHz<br>1 kHz to 10 kHz<br><br>200 mA to 2 A<br>10 Hz to 44 Hz<br>44 Hz to 1 kHz<br>1 kHz to 5 kHz<br><br>2 A to 30 A<br>10 Hz to 44 Hz<br>44 Hz to 100 Hz<br>100 Hz to 1 kHz | 0.23 % + 28 $\mu$ A<br>0.070 % + 5.1 $\mu$ A<br>0.58 % + 21 $\mu$ A<br><br>0.23 % + 270 $\mu$ A<br>0.070 % + 68 $\mu$ A<br>0.70 % + 200 $\mu$ A<br><br>0.23 % + 1.8 mA<br>0.11 % + 1.8 mA<br>0.70 % + 2.5 mA<br><br>0.23 % + 27 mA<br>0.11 % + 12 mA<br>0.35 % + 9.5 mA |  | P                |
| AC RESISTANCE<br>Measurement   | At 25 Hz and 75 Hz<br>0 $\Omega$ to 400 $\Omega$   | 0.0024 % + 100 $\mu\Omega$  | Measurement of resistors using AC bridge.  | P                |
| Calibration of 16 <sup>th</sup> /17 <sup>th</sup> Edition electrical testers<br>Loop Resistance (Generation) | 0 $\Omega$ to 10 $\Omega$<br>10 $\Omega$ to 100 $\Omega$<br>100 $\Omega$ to 1 k $\Omega$   | 0.58 % + 30 m $\Omega$<br>0.58 % + 36 m $\Omega$<br>0.58 % + 130 m $\Omega$   | Known 50 Hz resistance values for application to loop testers.                           |                  |
| Capacitance  | At 1 kHz:<br>1 nF<br>10 nF<br>20 nF<br>50 nF<br>100 nF<br>1 $\mu$ F<br>10 $\mu$ F  | 4.3 %<br>0.75 %<br>0.58 %<br>0.51 %<br>0.41 %<br>0.50 %<br>0.92 %   | Known, fixed values of capacitance for application to capacitance measuring instruments. | P                |
| FREQUENCY<br>Generation  | 10 Hz to 10 MHz  | 24 $\mu$ Hz/Hz  | Using synthesised source   | P                |
| Measurement  | 3 Hz to 5 Hz<br>5 Hz to 10 Hz<br>10 Hz to 40 Hz<br>40 Hz to 300 Hz<br>300 Hz to 3 kHz<br>3 kHz to 30 kHz<br>30 kHz to 300 kHz  | 9.5 mHz<br>6.8 mHz<br>15 mHz<br>38 mHz<br>360 mHz<br>3.6 Hz<br>35 Hz  | Measured at 1.0 V  | P                |



0461  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TMS Europe Ltd**  
Issue No: 060 Issue date: 30 May 2024

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge                              | Range  | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )                                  | Remarks   | Location<br>Code |
|---|--|---|---|------------------|
| Temperature indicators, calibration by electrical simulation          |  |   |   |                  |
| Base metal thermocouples  | -200 °C to 0 °C<br>0 °C to 1370 °C   | 0.20 °C<br>0.20 °C  | Including cold junction compensation  | P and S          |
| Noble metal thermocouples   | 0 °C to 600 °C<br>600 °C to 1800 °C  | 0.30 °C<br>0.30 °C  | Including cold junction compensation  | P and S          |
| Resistance thermometers<br>(Pt100)                                    | -200 °C to +800 °C   | 0.012 °C to 0.043 °C  | For 4-wire measurements.<br>The quoted uncertainties<br>may be increased for 2-<br>wire configurations.   | P and S          |
| <b>TIME INTERVAL</b>  |  |   |   |                  |
| Timers and stopwatches with<br>mechanical or electrical<br>triggering | 10 s to 24 hrs   | 0.003 0 s/h + 0.031 s   | Comparison with standard<br>timer. The uncertainties<br>quoted will be particularly<br>dependent on the<br>resolution and repeatability<br>of the device being<br>calibrated. | P and S          |
| Manual start count up or count<br>down timers.                        | 10 s to 24 hrs   | 0.004 s/h + 0.92 s  | Comparison with a<br>standard timer.  | P and S          |
| Optical Tachometers   | 18 rpm to 30 rpm<br>30 rpm to 60 rpm<br>60 rpm to 240 rpm<br>240 rpm to 600 rpm<br>600 rpm to 2400 rpm<br>2400 rpm to 18,000 rpm<br>18,000 rpm to 30,000 rpm<br>30,000 rpm to 90,000 rpm | 0.15 %<br>0.065 %<br>0.036 %<br>0.045 %<br>0.036 %<br>0.016 %<br>0.013 %<br>0.012 % | Application of optical<br>pulses of known repetition<br>rate.   | P                |



0461  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TMS Europe Ltd**  
Issue No: 060 Issue date: 30 May 2024

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge  | Range                               | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )             | Remarks   | Location<br>Code |
|---|-------------------------------------|--|---|------------------|
| HUMIDITY  |                                     |  |   |                  |
| Dew point   | -20 °C to +30 °C<br>30 °C to +46 °C | 0.20 °C<br>0.20 °C   | By comparison with dew-point hygrometer and Platinum Resistance Thermometers  | P                |
| Temperature probes in air   | 10 °C to 50 °C                      | 0.049 °C   | Instruments with an electrical output can be calibrated   |                  |
| Relative humidity   | Example conditions                  | Corresponding to above dew-point and temperature uncertainties | Humidity limits:<br>11 %rh to 95 %rh at 10 °C<br>5 %rh to 95 %rh at 20 °C<br>5 %rh to 95 %rh at 35 °C<br>5 %rh to 85 %rh at 50 °C |                  |
| At 10 °C  | 11 %rh<br>50 %rh<br>95 %rh          | 0.40 %rh<br>0.84 %rh<br>1.5 %rh                                |   |                  |
| At 20 °C  | 5 %rh<br>50 %rh<br>95 %rh           | 0.40 %rh<br>0.80 %rh<br>1.4 %rh                                |   |                  |
| At 50 °C  | 5 %rh<br>50 %rh<br>85 %rh           | 0.40 %rh<br>1.0 %rh<br>1.5 %rh                                 |   |                  |
| Temperature probes in air   | 10 °C to 50 °C                      | 0.40 °C to 0.90 °C   |   | S                |
| Temperature measurement in air  | 10 °C to 50 °C                      | 0.40 °C to 0.90 °C*  | *An additional uncertainty component will normally be applicable owing to the thermal variations within the test environment      |                  |
| Relative humidity probes and environmental controlled chambers inclusive of associated indicators controllers and recorders |                                     |  | By comparison with reference hygrometer and Platinum Resistance Thermometers  |                  |
| At 10 °C  | 10 %rh<br>50 %rh<br>90 %rh          | 1.1 %rh<br>2.3 %rh<br>3.7 %rh                                  |   |                  |
| At 20 °C  | 10 %rh<br>50 %rh<br>90 %rh          | 1.1 %rh<br>1.5 %rh<br>2.2 %rh                                  |   |                  |
| At 50 °C  | 10 %rh<br>50 %rh<br>90 %rh          | 1.2 %rh<br>2.2 %rh<br>3.0 %rh                                  |   |                  |



0461  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TMS Europe Ltd**  
Issue No: 060 Issue date: 30 May 2024

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )   | Remarks   | Location<br>Code |
|--|---|--|---|------------------|
| <b>PRESSURE</b>                          |   |  | Methods consistent with EURAMET CG17.   |                  |
| Gauge gas pressure                       | -2 kPa to 2 kPa   | 0.28 % + 1.9 Pa  | Instruments with an electrical output can be calibrated   | P and S          |
|  | -97 kPa to +2.0 MPa   | 0.0069 % + 55 Pa   |   | P                |
|  | -95 kPa to +2.0 MPa   | 0.0070 % + 0.35 kPa  |   | S                |
| Gauge gas absolute                       | 3 kPa to 2.1 MPa  | 0.0069 % + 59 Pa   | Achievable range may be reduced in the absence of a physical pressure port on the device  | P                |
|  | 75 kPa to 115 kPa   | 9.0 Pa   |   | P                |
|  | 75 kPa to 105 kPa   | 0.018 % + 58 Pa  |   | S                |
| Gauge hydraulic pressure                 | 5 kPa to 2.1 MPa  | 0.022 % + 0.40 kPa   |   | S                |
|  | 2 MPa to 70 MPa   | 0.018 % + 5.0 kPa  |   | P and S          |
| <b>TEMPERATURE</b>                       |   |  | Unless otherwise stated calibration performed by comparison with reference thermometers   |                  |
| Resistance Thermometers                  | -196 °C<br>-100 °C to -80 °C<br>-80 °C to 0 °C<br>Ice point 0 °C<br>0.01 °C<br>0 °C to 70 °C<br>70 °C to 300 °C<br>300 °C to 660 °C               | 0.010 °C<br>0.018 °C<br>0.011 °C<br>0.010 °C<br>0.0023 °C<br>0.020 °C<br>0.015 °C<br>0.19 °C | Liquid Nitrogen.<br>Metal Block Bath<br>Liquid Baths<br>Ice point<br>Triple point of water<br>Calibration performed within Liquid Baths<br>Calibration performed within Metal Block Baths | P                |
| Platinum thermocouples<br>R and S        | 0 °C to 70 °C<br>70 °C to 300 °C<br>300 °C to 650 °C<br>650 °C to 1000 °C<br>1000 °C to 1350 °C<br>1350 °C to 1600 °C                             | 0.46 °C<br>0.35 °C<br>0.36 °C<br>0.75 °C<br>1.1 °C<br>1.8 °C                                 | Calibration performed within Metal Block Baths or within Tube Furnaces  | P                |
| Platinum thermocouples<br>B              | 300 °C to 650 °C<br>650 °C to 1000 °C<br>1000 °C to 1350 °C<br>1350 °C to 1600 °C   | 0.58 °C<br>0.80 °C<br>1.1 °C<br>1.8 °C   |   |                  |
| Other thermocouples                      | -196 °C<br>-100 °C to -80 °C<br>-80 °C to 0 °C<br>0 °C to 70 °C<br>70 °C to 300 °C<br>300 °C to 650 °C<br>650 °C to 1000 °C<br>1000 °C to 1350 °C | 0.27 °C<br>0.17 °C<br>0.16 °C<br>0.15 °C<br>0.14 °C<br>0.24 °C<br>0.70 °C<br>1.0 °C          | Liquid Nitrogen.<br>Calibration performed within Metal Block Baths or within Tube Furnaces  | P                |



0461  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TMS Europe Ltd**

**Issue No: 060    Issue date: 30 May 2024**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge  | Range  | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )  | Remarks  | Location<br>Code |
|---|--|---|--|------------------|
| TEMPERATURE (continued)   |  |   |  |                  |
| Other thermocouples   | -196 °C<br>-100 °C to -80 °C<br>-80 °C to 0 °C<br>0 °C to 300 °C<br>300 °C to 1100 °C<br>1100 °C to 1350 °C  | 0.25 °C<br>0.30 °C<br>0.20 °C<br>0.20 °C<br>1.0 °C<br>2.5 °C  | Liquid Nitrogen.<br>Calibration performed<br>within block bath and Liquid<br>Baths or within Tube<br>Furnaces  | S                |
| Thermometers connected to<br>suitable indicators  | -196 °C<br>-100 °C to -80 °C<br>-80 °C to 0 °C<br>0 °C to 300 °C<br>300 °C to 400 °C   | 0.20 °C<br>0.30 °C<br>0.20 °C<br>0.10 °C<br>1.0 °C  | Including instruments<br>incorporated in customers<br>Freezers, fridges, ovens<br>chambers incubators and<br>furnaces  | S                |
| Temperature Controlled<br>Autoclaves, Chambers,<br>Environmental Cabinets,<br>Furnaces, Liquid Baths,<br>Fridges/Refrigerators, Freezers,<br>Incubators and Ovens | -80 °C to 0 °C<br>0 °C to 300 °C<br>300 °C to 1100 °C<br>1100 °C to 1600 °C  | 0.20 °C (PRTs)<br>0.50 °C (thermocouples)<br>0.10 °C (PRTs)<br>0.50 °C (thermocouples)<br>1.0 °C<br>2.5 °C                              | Single and multipoint time<br>dependent temperature<br>profiling, also referred to as<br>spatial temperature<br>surveying or mapping   | P and S          |
| Compensating and extension<br>cables  | -196 °C<br>-100 °C to -80 °C<br>-80 °C to +250 °C  | 0.25 °C<br>0.35 °C<br>0.25 °C   | Liquid Nitrogen.<br>Calibration performed<br>within Metal Block Baths or<br>within Tube Furnaces   | P and S          |
| Mechanical Dial type and<br>Electronic thermometers with<br>sensors   | Range as per sensor  | As for sensor   | Instruments with an<br>electrical output can be<br>calibrated  | P and S          |
| Metal block calibrators   | -100 °C to 0 °C<br>0 °C to 300 °C<br>300 °C to 660 °C<br>660 °C to 1100 °C   | 0.21 °C<br>0.093 °C<br>0.20 °C<br>0.95 °C   | Method consistent with<br>Euramet cg13   | P                |
| Metal block calibrators   | -50 °C to +300 °C<br>300 °C to 650 °C<br>650 °C to 1100 °C   | 0.21 °C<br>2.5 °C<br>4.0 °C   | Method consistent with<br>Euramet cg13   | S                |
| Radiation thermometers<br>(pyrometers)  | -50 °C to -30 °C<br>-30 °C to 0 °C<br>0 °C to +18 °C<br>18 °C to 35 °C<br>35 °C to 100 °C<br>100 °C to 200 °C<br>200 °C to 350 °C<br>350 °C to 480 °C<br>480 °C to 600 °C<br>600 °C to 900 °C<br>900 °C to 1200 °C<br>1200 °C to 1400 °C<br>1400 °C to 1500 °C | 1.3 °C<br>0.76 °C<br>0.26 °C<br>0.44 °C<br>0.53 °C<br>0.57 °C<br>0.74 °C<br>0.92 °C<br>0.98 °C<br>0.78 °C<br>1.0 °C<br>1.2 °C<br>1.4 °C | Calibration performed by<br>comparison with reference<br>radiation thermometer<br>For an emissivity of 1.0.<br>Other emissivities can be<br>specified but will attract an<br>additional uncertainty. | P                |

END





0461  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TMS Europe Ltd**  
**Issue No: 060**    **Issue date: 30 May 2024**

Calibration performed by the Organisation at the locations specified

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$