


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION</p> <p>0556</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Horiba MIRA Limited</p> <p>Issue No: 028 Issue date: 07 December 2021</p>	
	<p>Calibration Centre Watling Street Nuneaton Warwickshire CV10 0TU</p>	<p>Contact: Mr Philip Macleod Tel: +44 (0)2476 355643 Fax: +44 (0)2476 358225 E-Mail: calibration.centre@horiba-mira.com Website: www.horiba-mira.com</p>
<p>Calibration performed at the above address only</p>		

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
<p>PRESSURE</p> <p><u>Gas Pressure Gauge</u></p> <p>Calibration of pressure indicating instruments and gauges</p> <p><u>Gas Pressure Absolute</u></p> <p>Calibration of pressure indicating instruments and gauges</p> <p><u>Hydraulic Pressure Gauge</u></p> <p>Calibration of pressure indicating instruments and gauges</p>	<p>-90 kPa to 0 Pa 0 to 2 MPa</p> <p>10 kPa to 200 kPa 200 kPa to 2.1 MPa</p> <p>0 to 50 MPa</p>	<p>0.024 % + 0.59 kPa 0.024 % + 0.59 kPa</p> <p>0.15 kPa 0.022 % + 0.59 kPa</p> <p>0.016 % + 0.11 MPa</p>	<p>Methods consistent with EURAMET CG17.</p> <p>Calibration of pressure devices with an electrical output can be undertaken</p>
<p>Values and uncertainties listed below are applicable for the calibration of both measurement instruments and for instruments with an output. the method used is by direct comparison unless otherwise stated in the remarks column.</p>			
<p>ELECTRICAL</p> <p>DC VOLTAGE</p> <p>Generation</p> <p>Measurement</p>	<p>0 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V 330 V to 1000 V</p> <p>0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V</p>	<p>20 µV/V + 1.0 µV 11 µV/V + 2.0 µV 12 µV/V + 20 µV 18 µV/V + 150 µV 18 µV/V + 1.5 mV</p> <p>47 µV/V + 1.2 µV 31 µV/V + 4.0 µV 31 µV/V + 80 µV 47 µV/V + 600 µV 51 µV/V + 6.0 mV</p>	



0556

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Horiba MIRA Ltdimited

Issue No: 028 Issue date: 07 December 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
AC VOLTAGE Generation	1 mV to 33 mV 0.2 Hz to 10 Hz 10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 500 kHz 33 mV to 330 mV 0.2 Hz to 10 Hz 10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 500 kHz 0.33 V to 3.3 V 0.2 Hz to 10 Hz 10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 500 kHz 3.3 V to 10.23 V 0.2 Hz to 10 Hz 3.3 V to 33 V 10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 33 V to 330 V 10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 330 V to 1020 V 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	800 µV/V + 100 µV 800 µV/V + 100 µV 150 µV/V + 100 µV 200 µV/V + 100 µV 0.10 % + 100 µV 0.35 % + 110 µV 0.80 % + 150 µV 300 µV/V + 120 µV 300 µV/V + 120 µV 150 µV/V + 120 µV 160 µV/V + 120 µV 350 µV/V + 120 µV 800 µV/V + 140 µV 0.20 % + 190 µV 300 µV/V + 400 µV 300 µV/V + 400 µV 150 µV/V + 400 µV 190 µV/V + 400 µV 300 µV/V + 400 µV 700 µV/V + 460 µV 0.24 % + 1.2 mV 300 µV/V + 4.5 mV 300 µV/V + 4.0 mV 150 µV/V + 4.0 mV 240 µV/V + 4.0 mV 350 µV/V + 4.0 mV 900 µV/V + 5.0 mV 190 µV/V + 30 mV 200 µV/V + 30 mV 250 µV/V + 30 mV 300 µV/V + 30 mV 0.20 % + 80 mV 300 µV/V + 120 mV 250 µV/V + 120 mV 300 µV/V + 120 mV	



0556

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Horiba MIRA Ltdimited

Issue No: 028 Issue date: 07 December 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
AC VOLTAGE (cont'd) Measurement	20 mV to 200 mV 0.2 Hz to 10 Hz	0.40 %	
	1 mV to 200 mV 10 Hz to 50 Hz	0.30 % + 30 µV	
	50 Hz to 100 Hz	0.080 % + 30 µV	
	100 Hz to 30 kHz	0.060 % + 30 µV	
	30 kHz to 50 kHz	0.070 % + 30 µV	
	50 kHz to 100 kHz	0.36 % + 30 µV	
	100 kHz to 200 kHz	0.91 % + 50 µV	
	200 kHz to 1 MHz	2.0 % + 200 µV	
	1 MHz to 2 MHz	5.0 % + 400 µV	
	0.2 V to 2 V 0.2 Hz to 10 Hz	0.40 %	
	10 Hz to 50 Hz	0.31 % + 300 µV	
	50 Hz to 100 Hz	0.080 % + 300 µV	
100 Hz to 30 kHz	0.060 % + 300 µV		
30 kHz to 50 kHz	0.060 % + 300 µV		
50 kHz to 100 kHz	0.36 % + 300 µV		
100 kHz to 200 kHz	0.91 % + 500 µV		
200 kHz to 1 MHz	2.0 % + 2.0 mV		
1 MHz to 2 MHz	5.0 % + 4.0 mV		
2 V to 20 V 0.2 Hz to 10 Hz	0.40%		
10 Hz to 50 Hz	0.30 % + 3.0 mV		
50 Hz to 100 Hz	0.080 % + 3.0 mV		
100 Hz to 2 kHz	0.070 % + 3.0 mV		
2 kHz to 10 kHz	0.10 % + 3.0 mV		
30 kHz to 50 kHz	0.060 % + 30 µV		
50 kHz to 100 kHz	0.30 % + 300 µV		
10 kHz to 30 kHz	0.15 % + 3.0 mV		
30 kHz to 50 kHz	0.16 % + 3.0 mV		
50 kHz to 100 kHz	0.36 % + 3.0 mV		
100 kHz to 200 kHz	0.90 % + 5.0 mV		
200 kHz to 1 MHz	4.0 % + 40 mV		
1 MHz to 2 MHz	7.0 % + 40 mV		
20 V to 200 V 0.2 Hz to 10 Hz	0.40%		
10 Hz to 50 Hz	0.31 % + 30 mV		
50 Hz to 100 Hz	0.08 % + 30 mV		
100 Hz to 2 kHz	0.08 % + 30 mV		
2 kHz to 10 kHz	0.11 % + 30 mV		
10 kHz to 30 kHz	0.15 % + 30 mV		
30 kHz to 50 kHz	0.16 % + 30 mV		
50 kHz to 100 kHz	0.37 % + 30 mV		
100 kHz to 200 kHz	0.90 % + 5.0 mV		
200 kHz to 1 MHz	4.0 % + 40 mV		
1 MHz to 2 MHz	4.0 % + 40 mV		
200 V to 750 V 50 Hz to 100 Hz	0.39 % + 110 mV		
100 Hz to 2 kHz	0.19 % + 110 mV		
2 kHz to 10 kHz	0.23 % + 110 mV		
10 kHz to 25 kHz	0.28 % + 110 mV		



0556

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Horiba MIRA Ltdimited

Issue No: 028 Issue date: 07 December 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
DC CURRENT			
Generation	0 µA to 330 µA 330 µA to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA 330 mA to 1.1 A 1.1 A to 3 A 3 A to 11 A 11 A to 20.5 A	150 µA/A + 25 nA 110 µA/A + 100 nA 100 µA/A + 2.0 µA 100 µA/A + 20 µA 210 µA/A + 80 µA 390 µA/A + 1.0 mA 510 µA/A + 2.5 mA 0.10 % + 4.0 mA	For the calibration of toroidal clamp-on ammeters
	10 A to 16.5 A 16.5 A to 150 A 150 A to 1025 A	0.25 % + 3.0 mA 0.25 % + 1.0 A 0.25 % + 1.5 A	
Measurement	0 µA to 200 µA 200 µA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A	610 µA/A + 5.0 nA 480 µA/A + 40 nA 480 µA/A + 0.40 µA 600 µA/A + 4.0 µA 0.11 % + 40 µA	
AC CURRENT			
Generation	29 µA to 330 µA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 30 kHz	0.20 % + 730 nA 0.15 % + 700 nA 0.13 % + 700 nA 0.30 % + 700 nA 0.80 % + 870 nA 1.6 % + 2.0 µA	
	330 µA to 3.3 mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 30 kHz	0.20 % + 2.0 µA 0.13 % + 2.0 µA 0.10 % + 2.0 µA 0.20 % + 2.0 µA 0.50 % + 2.2 µA 1.0 % + 2.5 µA	
	3.3 mA to 33 mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 30 kHz	0.18 % + 10 µA 0.090 % + 10 µA 0.040 % + 10 µA 0.080 % + 10 µA 0.20 % + 15 µA 0.40 % + 15 µA	
	33 mA to 330 mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 30 kHz	0.18 % + 100 µA 0.090 % + 100 µA 0.040 % + 100 µA 0.10 % + 120 µA 0.20 % + 200 µA 0.40 % + 300 µA	



0556

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Horiba MIRA Ltdimited

Issue No: 028 Issue date: 07 December 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
AC CURRENT (cont'd)			
Generation (cont'd)	330 mA to 1.1 A 10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	0.18 % + 780 µA 0.050 % + 780 µA 0.60 % + 1.8 mA 2.5 % + 6.0 mA	
	1.1 A to 3 A 10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	0.18 % + 1.3 mA 0.060 % + 1.3 mA 0.60 % + 2.2 mA 2.5 % + 6.2 mA	
	3 A to 11 A 45 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 5 kHz	0.060 % + 9.0 mA 0.10 % + 9.0 mA 3.0 % + 9.0 mA	
	11 A to 20 A 45 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 5 kHz	0.12 % + 15 mA 0.15 % + 15 mA 3.0 % + 15 mA	
	10 A to 16.5 A 45 Hz to 65 Hz 65 Hz to 440 Hz	0.28 % + 3.0 mA 0.79 % + 3.0 mA	For the calibration of toroidal clamp-on ammeters
	16.5 A to 150 A 45 Hz to 65 Hz 65 Hz to 440 Hz	0.28 % + 13 mA 0.79 % + 13 mA	For the calibration of toroidal clamp-on ammeters
	150 A to 1025 A 45 Hz to 65 Hz 65 Hz to 440 Hz	0.28 % + 40 mA 0.79 % + 40 mA	For the calibration of toroidal clamp-on ammeters
Measurement	0 µA to 200 µA 40 Hz to 50 Hz 50 Hz to 200 Hz 200 Hz to 1 kHz 1 kHz to 10 kHz	0.42 % + 30 nA 0.24 % + 30 nA 0.48 % + 30 nA 0.57 % + 30 nA	
	200 µA to 2 mA 40 Hz to 50 Hz 50 Hz to 1 kHz 1 kHz Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 50 kHz 50 kHz to 100 kHz	0.36 % + 300 nA 0.18 % + 300 nA 0.15 % + 300 nA 0.15 % + 300 nA 0.30 % + 300 nA 0.50 % + 300 nA	
	2 mA to 20 mA 40 Hz to 50 Hz 50 Hz to 200 Hz 200 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 50 kHz 50 kHz to 100 kHz	0.36 % + 3.0 µA 0.18 % + 3.0 µA 0.15 % + 3.0 µA 0.15 % + 3.0 µA 0.25 % + 3.0 µA 0.30 % + 3.0 µA 0.50 % + 3.0 µA	



0556

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Horiba MIRA Ltdimited

Issue No: 028 Issue date: 07 December 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
AC CURRENT (cont'd) Measurement (cont'd)	20 mA to 200 mA 40 Hz to 50 Hz 50 Hz to 200 Hz 200 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 50 kHz 50 kHz to 100 kHz 0.2 A to 2 A 40 Hz to 50 Hz 50 Hz to 200 Hz 200 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 50 kHz	0.36 % + 30 µA 0.18 % + 30 µA 0.15 % + 30 µA 0.21 % + 30 µA 0.50 % + 30 µA 1.0 % + 30 µA 3.0 % + 30 µA 0.42 % + 300 µA 0.24 % + 300 µA 0.36 % + 300 µA 0.52 % + 300 µA 1.5 % + 300 µA 4.0 % + 300 µA	
DC RESISTANCE Generation	0 Ω to 11 Ω 11 Ω to 33 Ω 33 Ω to 110 Ω 110 Ω to 330 Ω 330 Ω to 1.1 kΩ 1.1 kΩ to 3.3 kΩ 3.3 kΩ to 11 kΩ 11 kΩ to 33 kΩ 33 kΩ to 110 kΩ 110 kΩ to 330 kΩ 330 kΩ to 1.1 MΩ 1.1 MΩ to 3.3 MΩ 3.3 MΩ to 11 MΩ 11 MΩ to 33 MΩ 33 MΩ to 110 MΩ 110 MΩ to 330 MΩ 330 MΩ to 1.1 GΩ	40 µΩ/Ω + 2.0 mΩ 30 µΩ/Ω + 3.1 mΩ 28 µΩ/Ω + 4.0 mΩ 28 µΩ/Ω + 6.3 mΩ 28 µΩ/Ω + 20 mΩ 28 µΩ/Ω + 51 mΩ 28 µΩ/Ω + 150 mΩ 28 µΩ/Ω + 480 mΩ 28 µΩ/Ω + 550 mΩ 32 µΩ/Ω + 14 Ω 32 µΩ/Ω + 24 Ω 60 µΩ/Ω + 110 Ω 130 µΩ/Ω + 360 Ω 250 µΩ/Ω + 4.1 kΩ 500 µΩ/Ω + 19 kΩ 0.30 % + 230 kΩ 1.5 % + 1.2 MΩ	
Measurement	0 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 kΩ 2 kΩ to 20 kΩ 20 kΩ to 200 kΩ 200 kΩ to 2 MΩ 2 MΩ to 20 MΩ 20 MΩ to 200 MΩ 200 MΩ to 1 GΩ	94 µΩ/Ω + 6.1 mΩ 67 µΩ/Ω + 7.4 mΩ 59 µΩ/Ω + 14 mΩ 69 µΩ/Ω + 80 mΩ 120 µΩ/Ω + 900 mΩ 190 µΩ/Ω + 9.0 Ω 0.11 % + 90 Ω 2.4 % + 20 kΩ 4.0 % + 200 kΩ	



0556

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Horiba MIRA Ltdimited

Issue No: 028 Issue date: 07 December 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
CALIBRATION OF CHARGE AMPLIFIERS	1 pC to 33 pC 200 mHz to 10 Hz 10 Hz to 45 Hz 63 Hz to 10 kHz	1.2 % 0.093 % + 0.0060 pC 0.020 % + 0.0060 pC	
	33 pC to 330 pC 200 mHz to 10 Hz 10 Hz to 45 Hz 63 Hz to 10 kHz	1.2 % 0.036 % + 0.0080 pC 0.020 % + 0.0080 pC	
	330 pC to 3300 pC 200 mHz to 10 Hz 10 Hz to 45 Hz 63 Hz to 10 kHz	1.2 % 0.036 % + 0.050 pC 0.020 % + 0.060 pC	
	3300 pC to 33000 pC 200 mHz to 10 Hz 10 Hz to 45 Hz 63 Hz to 10 kHz	1.2 % 0.036 % + 0.65 pC 0.020 % + 0.60 pC	
	33000 pC to 100000 pC 200 mHz to 10 Hz 10 Hz to 45 Hz 63 Hz to 10 kHz	1.2 % 0.024 % + 2.0 pC 0.025 % + 6.0 pC	
Attenuation With reference to a set value of 3 V	0 dB to 30 dB DC and 3 Hz to 10 kHz	0.020 dB	
TEMPERATURE SIMULATION Thermocouple type			
K	-200 °C to +1300 °C	0.30 °C	Including reference junction compensation
J	-210 °C to +1200 °C	0.26 °C	
N	-200 °C to +1300 °C	0.36 °C	
T	-250 °C to -150 °C	0.31 °C	Excluding reference junction compensation
T	-150 °C to +400 °C	0.25 °C	
K	-200 °C to +1300 °C	0.23 °C	
J	-210 °C to +1200 °C	0.16 °C	
N	-200 °C to +1300 °C	0.30 °C	
T	-250 °C to -150 °C	0.23 °C	
T	-150 °C to +400 °C	0.16 °C	



0556

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Horiba MIRA Ltdimited

Issue No: 028 Issue date: 07 December 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
FREQUENCY			
Generation	10 mHz to 30 MHz	2.0 in 10 ¹¹	
Measurement	1 mHz to 225 MHz	0.05 µHz/Hz	
ANGULAR VELOCITY			
Calibration of angular velocity sensors			Ambient temperature 18 °C to 22 °C.
Applied angular velocity	10 °/s to 2000 °/s	0.13 % of angular velocity	
Output voltage measurement	0 V to 200 mV 200 mV to 2 V 2 V to 20 V	560 µV/V + 150 µV 560 µV/V + 150 µV 560 µV/V + 170 µV	
END			



0556

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Horiba MIRA Ltdimited

Issue No: 028 Issue date: 07 December 2021

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$