


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0654 Accredited to ISO/IEC 17025:2017	Calibration, Maintenance and Repair Ltd	
	Issue No: 043 Issue date: 23 August 2024	
	11 Frensham Road Norwich NR3 2BT	Contact: Mr P K Clark Tel: +44 (0)1603 279557 Fax: +44 (0)1603 278008 E-Mail: support@cmrcalibrate.co.uk Website: www.cmrcalibrate.co.uk

Calibration performed by the Organisation at the locations specified

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address 11 Frensham Road Norwich NR3 2BT	Local contact: Mr P K Clark Tel: +44 (0)1603 279557	Electrical Calibration Accelerometer calibration Mass Calibration NAWI Calibration	A
Address 5 Octavian Way, Team Valley Trading Est Gateshead NE11 0HZ	Local contact: Mr John Fryer Tel: +44 (0)191 4875951	Electrical Calibration Pressure Calibration	B



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
ELECTRICAL MEASUREMENTS			Calibrations are performed as a direct comparison against a reference standard	A
DC VOLTAGE				
Generation	0 mV to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1100 V	0.55 μ V 1.4 μ V 6.6 μ V 44 μ V 630 μ V 8.0 mV	For the calibration of measuring instruments	
	1 kV to 2 kV 2 kV to 3 kV 3 kV to 4 kV 4 kV to 5 kV	0.45 % + 7.0 V 0.45 % + 8.0 V 0.45 % + 9.0 V 0.45 % + 11 V		
Measurement	0 mV to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1 kV	1.0 μ V 2.0 μ V 10 μ V 98 μ V 1.3 mV 14 mV	For instruments with a voltage output	
	1 kV to 2 kV 2 kV to 3 kV 3 kV to 4 kV 4 kV to 5 kV	0.67 % + 9.0 V 0.67 % + 11 V 0.67 % + 14 V 0.67 % + 16 V		
DC RESISTANCE				A
Sourcing - Spot Values	1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 k Ω 1.9 k Ω 10 k Ω 19 k Ω 100 k Ω 190 k Ω 1 M Ω 1.9 M Ω 10 M Ω 19 M Ω 100 M Ω	110 μ Ω 210 μ Ω 270 μ Ω 540 μ Ω 1.2 m Ω 2.5 m Ω 8.7 m Ω 17 m Ω 81 m Ω 170 m Ω 1.1 Ω 2.4 Ω 16 Ω 48 Ω 460 Ω 1.1 k Ω 12 k Ω	For the calibration of measuring instruments	



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
DC RESISTANCE (continued)				A
Sourcing - Range values	0Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 33 MΩ 33 MΩ to 110 MΩ 110 MΩ to 330 MΩ 330 MΩ to 1.1 GΩ	2.0 mΩ 5.0 mΩ 31 mΩ 310 mΩ 3.0 Ω 37 Ω 2.0 kΩ 11 kΩ 60 kΩ 1.0 MΩ 17 MΩ	For the calibration of measuring instruments	
Measurement	0 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 GΩ 10 GΩ to 100 GΩ	250 μΩ 2.0 mΩ 14 mΩ 140 mΩ 1.0 Ω 22 Ω 0.71 % 1.0 %	For instruments with a resistance output	
DC CURRENT				A
Generation	0 μA to 10 μA 10 μA to 100 μA 100 μA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 3 A 3 A to 5 A 5 A to 10 A 10 A to 20 A 20 A to 30 A 16 A to 160 A 160 A to 1000 A	7.4 nA 12 nA 49 nA 450 nA 6.0 μA 110 μA 1.0 mA 4.0 mA 6.0 mA 12 mA 570 mA 25 mA 580 mA	For the calibration of measuring instruments Current clamp calibration using a multi turn coil	



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
DC CURRENT (continued)				A
Measurement	0 μ A to 10 μ A 10 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 2 A 2 A to 10 A 10 A to 20 A 20 A to 30 A	2.0 nA 4.0 nA 34 nA 340 nA 5.0 μ A 150 μ A 2.4 mA 25 mA 330 mA 490 mA	For instruments which generate current	A
AC VOLTAGE				A
Generation	1 Hz to 10 Hz 10 μ V to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 7 V 10 μV to 2.2 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40Hz to 500 Hz 500 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz 2.2 mV to 22 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40Hz to 500 Hz 500 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	60 μ V 270 μ V 1.9 mV 9.2 mV 5.3 μ V 4.9 μ V 4.9 μ V 4.9 μ V 4.9 μ V 4.9 μ V 5.2 μ V 7.1 μ V 14 μ V 27 μ V 30 μ V 11 μ V 7.1 μ V 6.9 μ V 6.9 μ V 6.9 μ V 6.9 μ V 10 μ V 19 μ V 38 μ V 59 μ V 92 μ V	For the calibration of measuring instruments	A



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
<p>Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.</p>				
AC VOLTAGE (continued)				A
Generation (continued)	<p>22 mV to 220 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40Hz to 500 Hz 500 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz</p> <p>220 mV to 2.2 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40Hz to 500 Hz 500 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz</p> <p>2.2 V to 22 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40Hz to 500 Hz 500 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz</p> <p>22 V to 220 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40Hz to 500 Hz 500 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz</p>	<p>76 μV 31 μV 23 μV 23 μV 23 μV 23 μV 39 μV 99 μV 190 μV 380 μV 740 μV</p> <p>660 μV 250 μV 120 μV 120 μV 120 μV 120 μV 180 μV 250 μV 950 μV 2.8 mV 4.7 mV</p> <p>6.6 mV 2.5 mV 1.1 mV 1.1 mV 1.1 mV 1.1 mV 1.8 mV 2.3 mV 7.2 mV 28 mV 42 mV</p> <p>66 mV 25 mV 14 mV 14 mV 14 mV 14 mV 22 mV 41 mV</p>	For measuring instruments	



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd

Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
AC VOLTAGE (continued)				A
Generation (continued)	220 V to 1100 V 55 Hz to 1 kHz 220 V to 500 V 50 Hz to 10 kHz 500 V to 750 V 50 Hz to 10 kHz 750 V to 1 kV 50 Hz to 10 kHz 1 kV to 2 kV 50 Hz to 60 Hz 2 kV to 3 kV 50 Hz to 60 Hz 3 kV to 4 kV 50 Hz to 60 Hz 4 kV to 5 kV 50 Hz to 60 Hz	110 mV 180 mV 260 mV 350 mV 0.67 % + 12 V 0.67 % + 16 V 0.67 % + 20 V 0.67 % + 24 V	For measuring instruments	
Measurement	30 µV to 10 mV 1 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 1 MHz 10 mV to 100 mV 1 Hz to 40 Hz 40 Hz to 1 kHz 1kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 100mV to 1V 1 Hz to 40 Hz 40 Hz to 1 kHz 1kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 1 V to 10 V 1 Hz to 40 Hz 40 Hz to 1 kHz 1kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	8.0 µV 5.0 µV 6.0 µV 13 µV 60 µV 160 µV 15 µV 13 µV 20 µV 38 µV 100 µV 360 µV 2.3 mV 140 µV 120 µV 200 µV 390 µV 970 µV 3.6 mV 24 mV 2.0 mV 1.0 mV 2.0 mV 4.0 mV 10 mV 36 mV 240 mV	For instruments with a voltage output	



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd

Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
<p>Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.</p>				
AC VOLTAGE (continued) Measurement (continued)	10V to 100V 1 Hz to 40 Hz 40 Hz to 1kHz 1kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100kHz to 300 kHz	38 mV 36 mV 37 mV 51 mV 150 mV 480 mV	For instruments with a voltage output.	A
	100V to 1 kV 1 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	520 mV 490 mV 720 mV 1.4 V 3.5 V		
AC CURRENT Generation	1 kV to 2 kV 30 Hz to 200 Hz 2 kV to 3 kV 30 Hz to 200 Hz 3 kV to 4 kV 30 Hz to 200 Hz 4 kV to 5 kV 30 Hz to 200 Hz	0.45 % + 8.0 V 0.45 % + 11 V 0.45 % + 13 V 0.45 % + 16 V	For the calibration of measuring instruments	A
	10 µA to 220 µA 55 Hz to 400 Hz 400 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	38 nA 38 nA 88 nA 360 nA		
	220 µA to 2.2 mA 55 Hz to 400 Hz 400 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	320 nA 320 nA 660 nA 3.6 µA		
	2.2 mA to 22 mA 55 Hz to 400 Hz 400 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	29 µA 29 µA 55 µA 290 µA		
	22 mA to 220 mA 55 Hz to 400 Hz 400 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	3.0 µA 3.0 µA 5.7 µA 34 µA		
	220 mA to 2.2 A 55 Hz to 400 Hz 400 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	660 µA 660 µA 1.3 mA 18 mA		



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
AC CURRENT (continued) Measurement (continued)	1 A to 10 A 50 Hz to 400 Hz 400 Hz to 1 kHz 10 A to 20 A 50 Hz to 400 Hz 400 Hz to 1 kHz 20 A to 30 A 50 Hz to 400 Hz 400 Hz to 1 kHz 30 A to 40 A 50 Hz to 400 Hz 400 Hz to 1 kHz 40 A to 50 A 50 Hz to 400 Hz 400 Hz to 1 kHz	 180 mA 140 mA 270 mA 250 mA 370 mA 360 mA 480 mA 470 mA 590 mA 580 mA		A
FREQUENCY Generation	0.5 Hz to 5.4 GHz	2.1 in 10^9	For the calibration of measuring instruments May be reported as events per unit time	A
Measurement	1 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 27 GHz 0.5 Hz to 5.4 GHz 5.4 GHz to 20 GHz	35 μ Hz 0.40 Hz 0.24 μ Hz/Hz 2.1 in 10^9 4.0 MHz	For instruments outputting frequency For the calibration of instruments with frequency outputs. May be reported as events per unit time	A
TIME and FREQUENCY Elapsed time			Mechanical timers / stop watches	A
Single event	10 s to 12 hrs	0.23 s		
Revolutions Per Minute	60 RPM to 1000 RPM 1000 RPM to 30000 RPM 30000 RPM to 96000 RPM	0.37 RPM 1.6 RPM 23 RPM		



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
CAPACITANCE				A
Measurement For capacitance sources	1 kHz 0.1 pF to 1 pF 1 pF to 10 pF 10 pF to 100 pF 100 pF to 1000 pF 1 MHz 0.1 pF to 1 pF 1 pF to 10 pF 10 pF to 100 pF 100 pF to 1000 pF	0.030 pF 0.035 pF 0.10 pF 1.2 pF 0.0050 pF 0.030 pF 0.90 pF 8.0 pF	For capacitance sources	
Simulated For capacitance meters	10 Hz to 10 kHz 220 pF to 400 pF 400 pF to 1.1 nF 10 Hz to 3 kHz 1.1 nF to 3.3 nF 10 Hz to 1 kHz 3.3 nF to 11 nF 11 nF to 33 nF 10 Hz to 1 kHz 33 nF to 110 nF 110 nF to 330 nF 10 Hz to 600 Hz 330 nF to 1.1 μF 10 Hz to 300 Hz 1.1 μF to 3.3 μF 10 Hz to 150 Hz 3.3 μF to 11 μF 10 Hz to 120 Hz 11 μF to 33 μF 10 Hz to 80 Hz 33 μF to 110 μF	13 pF 16 pF 27 pF 39 pF 180 pF 380 pF 1.0 nF 5.0 nF 12 nF 39 nF 140 nF 610 nF		



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
CAPACITANCE (continued) For capacitance meters	0 Hz to 50 Hz 110 μ F to 330 μ F	2.0 μ F		A
	0 Hz to 20 Hz 330 μ F to 1.1 mF	6.0 μ F		
	0 Hz to 6 Hz 1.1 mF to 3.3 mF	2.0 μ F		
	0 Hz to 2 Hz 3.3 mF to 11 mF	60 μ F		
	0 Hz to 0.6 Hz 11 mF to 33 mF	280 μ F		
	0 Hz to 0.2 Hz 33 mF to 110 mF	1.0 mF		
CALIBRATION OF INSULATION TESTERS				A
Insulation Resistance	100 k Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω	2.0 k Ω 32 k Ω 510 k Ω 7.0 M Ω	Up to 1350 V	
Insulation test voltage Nominal	50 V 100 V 250 V 500 V 1 kV	1.0 V 2.0 V 3.0 V 4.0 V 7.0 V	The test voltage will normally be measured with a 1 mA load.	
Continuity Resistance	0 Ω to 1 Ω 1 Ω to 50 Ω 50 Ω to 100 Ω 100 Ω to 200 Ω	58 m Ω 100 m Ω 120 m Ω 140 m Ω		



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
Temperature indicators, calibration by electrical simulation				A
Type K thermocouple	-200 °C to +1372 °C	0.73 °C	Including reference junction compensation	
Type T thermocouple	-250 °C to +400 °C	0.65 °C	Including reference junction compensation	
Type N thermocouple	-200 °C to +1300 °C	0.46 °C	Including reference junction compensation	
Type S thermocouple	0 °C to 1767 °C	0.77 °C	Including reference junction compensation	
Type J thermocouple	-210 °C to +1200 °C	0.23 °C	Including reference junction compensation	
Type E thermocouple	-250 °C to +1000 °C	0.26 °C	Including reference junction compensation	
Resistance thermometer (Pt 100)	-200°C to +800°C	0.05°C		
Calibration of Oscilloscopes				A
Voltage deflection	0 V to 6 V 11 V to 130 V	0.33 % 0.22 %	Into 50 Ω Into 1 MΩ	
Edge verification - Amplitude	4.5 mV to 2.27 V 1 kHz, 10 kHz & 100 kHz	2.7 %		
Time markers	100 ns to 20 ms 20 ms to 50 ms	83 ns 250 μs		
Rise time	1 kHz to 2 MHz 2 MHz to 10 MHz	300 ps 350 ps		
Bandwidth Level flatness 50 kHz ref Nominal 3 V	To 100 MHz To 300 MHz To 600 MHz To 1.1 GHz	3.3 % 5.4 % 8.5 % 9.5 %	CMC is with respect to relative amplitude of level flatness this will be quoted in frequency terms when stating bandwidth uncertainty as it will vary from device to device.	



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
ACCELEROMETRY			Uncertainties for accelerometry sensitivity calibrations at ambient temperature applies for masses up to 40 grams.	A
Accelerometer calibration Charge	Sensitivity range: 0.3 pC/g to 1000 pC/g (0.03 pC/ms ² to 100 pC/ms ²) 1 Hz to 2 Hz 2 Hz to 5 Hz 5 Hz to 5 kHz 5 kHz to 10 kHz	3.0 % 1.2 % 0.80 % 0.90 %	Nominal acceleration ms ⁻² : 1 Hz: 0.10 3 Hz: 0.20 4 Hz to 10 kHz: 0.3 to 100.0	
Devices with integral electronics	Sensitivity range: 0.04 pC/g to 0.3 pC/g (0.004 pC/ms ² to 0.03 pC/ms ²) 20 Hz to 10 kHz	3.0 %	Nominal acceleration 0.3 ms ⁻² to 100.0 ms ⁻² :	
Voltage (mV/ms ²)	0.03 mV/ms ² to 1000 mV/ms ² 1 Hz to 2 Hz 2 Hz to 5 Hz 5 Hz to 5 kHz 5 kHz to 10 kHz	2.0 % 0.83 % 0.80 % 0.90 %	Nominal acceleration ms ⁻² : 1 Hz: 0.10 3 Hz: 0.20 4 Hz to 10 kHz: 0.3 to 100.0	
Resistive/Capacitive	0.004 mV/ms ² to 0.03 mV/ms ² 20 Hz to 10 kHz	3.0 %		
	0.01 mV/ms ² to 100 mV/ms ² 1 Hz to 2 Hz 2 Hz to 5 Hz 5 Hz to 5 kHz 5 kHz to 10 kHz	2.0 % 0.83 % 0.86 % 0.90 %	Nominal acceleration ms ⁻² : 1 Hz: 0.10 3 Hz: 0.20 4 Hz to 10 kHz: 0.3 to 100.0	
Vibration	10 Hz to 5 kHz		Reference to transfer standard Accelerometer	
Acceleration	0.1g to 10 g	0.40 %		
Velocity	3 mm/s to 2000 mm/s	0.40 %		
Displacement	0.2µm to 100 mm	0.40 %		
Velocity transducers	5 Hz to 10 kHz 1.5 mm/s to 3000 mm/s	0.90 %		



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
<p>Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.</p>				
<p>ACCELEROMETRY (continued) Transducer at any temperature from -60 °C to +180 °C:</p>	<p>0.03 mV/ms² to 1000 mV/ms² 20 Hz to 630 Hz</p>	<p>3.0 %</p>	<p>At a nominal 1 g_n This uncertainty applies for masses up to 60 grams, larger masses can be calibrated but at greater uncertainties.</p>	A
<p>Portable calibrators</p>	<p>Frequency 160 Hz 10 ms² nominal</p>	<p>1.5 %</p>	<p>Note indicated outputs on Bench Calibrators for acceleration are normally indicated in "g".</p>	
<p>Bench Calibrators</p>	<p>10mV to 10V, 5 Hz to 10KHz Up to 25g</p>	<p>2.4 %</p>		
<p>Acceleration</p>		<p>2.4 %</p>		
<p>Velocity</p>		<p>2.4 %</p>		
<p>Displacement</p>		<p>2.4 %</p>		
<p>DISTORTION THD</p>		<p>20 Hz to 20 kHz 10 mV to 750 V 0 % to 1 % 1 % to 100 %</p>		



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
ACCELERATION TRANSDUCERS - SHOCK CALIBRATION				A
<u>Piezoelectric and Piezoresistive and voltage types</u>				
Transducer at nominal 23 °C	Sensitivity 0.001 pC/ms ² to 100 pC/ms ² 0.001 mV/ms ² to 10 mV/ms ² Within the range of: 2 ms ² to 1000 ms ²	3.0 %	Calibration of charge sensitivity by comparison with a reference (precision grade) transducer The transducer to be calibrated must have a mass of no more than 40 grams.	
Vibration meters and analysers	Frequency 10 Hz to 1 kHz Range 0.1 ms ² to 1 ms ²	2.0 %		
Shock and Impulse hammers	Half Sine Wave Frequency 500 Hz to 20 kHz Applied Force 0.2 N to 450 N	4.0 %		
Spring hammers				
Impact Energy Imparted from Spring Operated Impact Test Apparatus - as specified in BS EN 60068-2-75:1997 and IEC 60068-2-75 1997	0.1 Joule to 1.0 Joule	0.015 Joule	Calibrations can be given in Joule or Newton Meter units.	
HELICOPTER BLADE BALANCING TEST SETS				
Strobe frequency	1 Hz to 10 Hz 20 Hz to 50 Hz 50 Hz to 100 Hz 100 Hz to 150 Hz 150 Hz to 200 Hz	1.2 Hz 1.2 Hz 1.4 Hz 1.6 Hz 1.7 Hz		
MASS	500 g 200 g 100 g	9 mg 7 mg 7 mg	Calibrated using Borda substitution method.	A
NON-AUTOMATIC WEIGHING MACHINES	100 g 200 g 500 g	19 mg 19 mg 20 mg	Weights are available in OIML Class: M1 from 1 g to 500 g Calibration method in line with the requirements of Euramet gauge cg-18	



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
Gateshead Laboratory				
ELECTRICAL				
DC VOLTAGE Generation	0 mV to 50 mV 50 mV to 100 mV 100 mV to 220 mV 0.22 V to 0.5 V 0.5 to 1 V 1 V to 2.2 V 2.2 V to 5 V 5 V to 11 V 11 V to 22 V 22 V to 50 V 50 V to 100 V 100 V to 220 V 220 V to 500 V 500 V to 1100 V	1.1 μ V 1.4 μ V 2.5 μ V 4.9 μ V 9.4 μ V 20 μ V 47 μ V 0.10 mV 0.20 mV 0.54 mV 1.0 mV 2.3 mV 6.5 mV 14 mV		B
Measurement	0 mV 0 mV to 50 mV 50 mV to 100 mV 100 mV to 200 mV 200 mV to 500 mV 500 mV to 1 V 1 V to 2 V 2 V to 5 V 5 V to 10 V 10 V to 20 V 20 V to 50 V 50 V to 100 V 100 V to 200 V 200 V to 500 V 500 V to 1050 V	0.65 μ V 0.86 μ V 1.1 μ V 2.0 μ V 2.1 μ V 3.6 μ V 7.0 μ V 18 μ V 34 μ V 69 μ V 0.30 mV 0.53 mV 1.1 mV 2.8 mV 5.5 mV		



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
DC RESISTANCE Sourcing - Spot Values	0 Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 k Ω 1.9 k Ω 10 k Ω 19 k Ω 100 k Ω 190 k Ω 1 M Ω 1.9 M Ω 10 M Ω 19 M Ω 100 M Ω	58 $\mu\Omega$ 0.13 m Ω 0.24 m Ω 0.38 m Ω 0.68 m Ω 2.3 m Ω 4.4 m Ω 17 m Ω 33 m Ω 0.16 Ω 0.31 Ω 1.8 Ω 3.5 Ω 27 Ω 53 Ω 0.53 k Ω 1.2 k Ω 15 k Ω		B
DC RESISTANCE Sourcing - Range values	0 Ω to 11 Ω 11 Ω to 33 Ω 33 Ω to 110 Ω 110 Ω to 330 Ω 330 Ω to 1.1 k Ω 1.1 k Ω to 3.3 k Ω 3.3 k Ω to 11 k Ω 11 k Ω to 33 k Ω 33 k Ω to 110 k Ω 110 k Ω to 330 k Ω 330 k Ω to 1.1 M Ω 1.1 M Ω to 3.3 M Ω 3.3 M Ω to 11 M Ω 11 M Ω to 33 M Ω 33 M Ω to 110 M Ω 110 M Ω to 330 M Ω 330 M Ω to 1.1 G Ω	0.67 m Ω 1.1 m Ω 3.6 m Ω 11 m Ω 36 m Ω 110 m Ω 0.36 m Ω 1.1 Ω 3.6 Ω 12 Ω 41 Ω 230 Ω 1.7 k Ω 9.6 k Ω 64 k Ω 1.1 M Ω 19 M Ω		B



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd

Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
<p>Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.</p>				
DC RESISTANCE Measurement	0 Ω 0 Ω to 0.5 Ω 0.5 Ω to 1 Ω 1 Ω to 2 Ω 2 Ω to 5 Ω 5 Ω to 10 Ω 10 Ω to 20 Ω 20 Ω to 50 Ω 50 Ω to 100 Ω 100 Ω to 200 Ω 0.2 k Ω to 0.5 k Ω 0.5 k Ω to 1 k Ω 1 k Ω to 2 k Ω 2 k Ω to 5 k Ω 5 k Ω to 10 k Ω 10 k Ω to 20 k Ω 20 k Ω to 50 k Ω 50 k Ω to 100 k Ω 100 k Ω to 200 k Ω 0.2 M Ω to 0.5 M Ω 0.5 M Ω to 1 M Ω 1 M Ω to 2 M Ω 2 M Ω to 5 M Ω 5 M Ω to 10 M Ω 10 M Ω to 20 M Ω 20 M Ω to 50 M Ω 50 M Ω to 100 M Ω 100 M Ω to 1 G Ω	7.5 $\mu\Omega$ 13 $\mu\Omega$ 21 $\mu\Omega$ 43 $\mu\Omega$ 92 $\mu\Omega$ 0.14 m Ω 0.25 m Ω 0.58 m Ω 1.1 m Ω 2.2 m Ω 5.5 m Ω 11 m Ω 21 m Ω 57 m Ω 0.11 Ω 0.22 Ω 0.62 Ω 1.1 Ω 2.2 Ω 7.8 Ω 13 Ω 26 Ω 0.18 k Ω 0.26 k Ω 0.52 k Ω 14 k Ω 19 k Ω 0.34 M Ω		B
DC CURRENT Generation	0 μA to 50 μA 50 μA to 100 μA 100 μA to 220 μA 0.22 mA to 0.5 mA 0.5 mA to 1 mA 1 mA to 2.2 mA 2.2 mA to 5 mA 5 mA to 10 mA 10 mA to 22 mA 22 mA to 50 mA 50 mA to 100 mA 100 mA to 220 mA 0.22 A to 0.5 A 0.5 A to 1 A 1 A to 2.2 A 2.2 A to 3 A 3 A to 11 A 11 A to 20.5 A	12 nA 13 nA 19 nA 37 nA 71 nA 0.15 μA 0.37 μA 0.71 μA 1.5 μA 4.2 μA 8.2 μA 18 μA 65 μA 0.12 mA 0.24 mA 1.3 mA 6.4 mA 24 mA		B



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
DC CURRENT (continued) Generation	20 A to 30 A 30 A to 110 A 110 A to 200 A 200 A to 550 A 550 A to 1025 A	150 mA 500 mA 930 mA 2.5 A 4.7 A	Current clamp calibration using a multi turn coil	B
Measurement	0 μ A to 5 μ A 5 μ A to 10 μ A 10 μ A to 20 μ A 20 μ A to 50 μ A 50 μ A to 100 μ A 100 μ A to 200 μ A 0.2 mA to 0.5 mA 0.5 mA to 1 mA 1 mA to 2 mA 2 mA to 5 mA 5 mA to 10 mA 10 mA to 20 mA 20 mA to 50 mA 50 mA to 100 mA 100 mA to 200 mA 0.2 A to 0.5 A 0.5 A to 1 A 1 A to 2 A 2 A to 5 A 5 A to 10 A 10 A to 20 A 20 A to 30 A	0.57 nA 0.63 nA 1.2 nA 1.1 nA 1.4 nA 2.9 nA 9.1 nA 13 nA 26 nA 0.11 μ A 0.18 μ A 0.35 μ A 3.7 μ A 6.8 μ A 14 μ A 0.14 mA 0.19 mA 0.39 mA 1.5 mA 2.8 mA 14 mA 20 mA		
AC VOLTAGE Generation	0.22 mV to 0.5 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz	5.8 μ V 5.8 μ V 5.8 μ V 5.8 μ V 5.8 μ V 5.8 μ V 5.8 μ V 9.3 μ V 17 μ V 35 μ V 46 μ V		B



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd

Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
AC VOLTAGE Generation (continued)	<p>0.5 mV to 1 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz</p> <p>1 mV to 2.2 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz</p> <p>2.2 mV to 5 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz</p> <p>5 mV to 10 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz</p>	<p>5.9 μV 5.8 μV 5.8 μV 5.8 μV 5.8 μV 5.8 μV 5.8 μV 5.8 μV 9.3 μV 17 μV 35 μV 47 μV</p> <p>6.0 μV 5.8 μV 5.8 μV 5.8 μV 5.8 μV 5.8 μV 5.9 μV 10 μV 18 μV 35 μV 48 μV</p> <p>7.8 μV 7.1 μV 7.0 μV 7.0 μV 7.0 μV 7.0 μV 7.4 μV 11 μV 19 μV 36 μV 54 μV</p> <p>10 μV 7.5 μV 7.1 μV 7.1 μV 7.1 μV 7.1 μV 8.4 μV 14 μV 23 μV 40 μV 72 μV</p>		B



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd

Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
AC VOLTAGE Generation (continued)	<p>10 mV to 22 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz</p> <p>22 mV to 50 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz</p> <p>50 mV to 100 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz</p> <p>100 mV to 220 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz</p>	<p>17 μV 9.3 μV 7.6 μV 7.6 μV 7.6 μV 7.6 μV 13 μV 26 μV 37 μV 57 μV 0.13 mV</p> <p>39 μV 18 μV 13 μV 13 μV 13 μV 13 μV 24 μV 62 μV 72 μV 0.11 mV 0.24 mV</p> <p>72 μV 30 μV 17 μV 17 μV 17 μV 17 μV 43 μV 0.11 mV 0.13 mV 0.21 mV 0.43 mV</p> <p>0.15 mV 62 μV 30 μV 30 μV 30 μV 30 μV 92 μV 0.23 mV 0.28 mV 0.46 mV 0.92 mV</p>		B



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd

Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
AC VOLTAGE Generation (continued)	<p>220 mV to 0.5 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz</p> <p>0.5 V to 1 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz</p> <p>1 V to 2.2 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz</p> <p>2.2 V to 5 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz</p>	<p>0.37 mV 0.11 mV 52 μV 52 μV 52 μV 52 μV 85 μV 0.19 mV 0.33 mV 0.83 mV 1.8 mV</p> <p>0.70 mV 0.21 mV 0.10 mV 0.10 mV 0.10 mV 0.10 mV 0.16 mV 0.34 mV 0.58 mV 1.5 mV 3.0 mV</p> <p>1.5 mV 0.46 mV 0.22 mV 0.22 mV 0.22 mV 0.22 mV 0.36 mV 0.72 mV 1.2 mV 3.1 mV 6.2 mV</p> <p>3.7 mV 1.1 mV 0.52 mV 0.52 mV 0.52 mV 0.52 mV 0.85 mV 1.7 mV 4.0 mV 10 mV 20 mV</p>		B



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
AC VOLTAGE Generation (continued)	5 V to 10 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz 10 V to 22 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz 22 V to 50 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 50 V to 100 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	7.0 mV 2.1 mV 1.0 mV 1.0 mV 1.0 mV 1.0 mV 1.6 mV 3.3 mV 7.2 mV 17 mV 36 mV 15 mV 4.6 mV 2.2 mV 2.2 mV 2.2 mV 2.2 mV 3.6 mV 7.1 mV 15 mV 36 mV 77 mV 37 mV 11 mV 5.8 mV 5.8 mV 5.8 mV 5.8 mV 15 mV 37 mV 70 mV 21 mV 11 mV 11 mV 11 mV 11 mV 29 mV 70 mV		B



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
AC VOLTAGE Generation (continued)	100 V to 220 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 220 V to 500 V 50 Hz 50 Hz to 500 Hz 500 Hz to 1 kHz 500 V to 1100 V 50 Hz 50 Hz to 500 Hz 500 Hz to 1 kHz	0.15 V 46 mV 23 mV 23 mV 23 mV 23 mV 64 mV 0.15 mV 57 mV 57 mV 57 mV 0.12 V 0.12 V 0.12 V		B
AC VOLTAGE Measurement	1 mV to 5 mV 10 Hz to 2 kHz 2 kHz to 10 kHz. 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 5 mV to 10 mV 10 Hz 10 Hz to 2 kHz 2 kHz to 10 kHz. 10 kHz to 30 kHz 30 kHz to 60 kHz 60 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	3.0 μ V 3.3 μ V 3.3 μ V 18 μ V 58 μ V 0.12 mV 3.0 μ V 4.2 μ V 4.9 μ V 5.0 μ V 36 μ V 35 μ V 0.12 mV 0.23 mV		B



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
AC VOLTAGE Measurement (continued)	<p>10 mV to 50 mV 10Hz 10 Hz to 2 kHz 2 kHz to 10 kHz. 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 1 MHz to 2 MHz 2 MHz to 4 MHz 4 MHz to 6 MHz 6 MHz to 8 MHz 8 MHz to 10 MHz</p> <p>50 mV to 100 mV 10 Hz 10 Hz to 2 kHz 2 kHz to 10 kHz. 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 1 MHz to 2 MHz 2 MHz to 4 MHz 4 MHz to 8 MHz 8 MHz to 10 MHz</p> <p>0.1 V to 0.5 V 10Hz 10 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 1 MHz to 2 MHz 2 MHz to 4 MHz 4 MHz to 8MHz 8 MHz to 10 MHz</p> <p>0.5 V to 1 V 10Hz 10 Hz to 2 kHz 2 kHz to 10 kHz. 10 kHz to 30 kHz 30 kHz to 60 kHz 60 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 1 MHz to 2 MHz 2 MHz to 4 MHz 4 MHz to 8 MHz 8 MHz to 10 MHz</p>	<p>6.1 μV 5.7 μV 7.9 μV 14 μV 34 μV 0.14 mV 0.69 mV 1.1 mV 2.7 mV 5.1 mV 5.2 mV 9.5 mV</p> <p>11 10 μV 15 μV 27 μV 68 μV 0.25 mV 1.3 mV 1.9 mV 4.9 mV 10 mV 19 mV</p> <p>50 μV 47 μV 73 μV 0.13 mV 0.34 mV 1.4 mV 6.3 mV 11 mV 26 mV 51 mV 92 mV</p> <p>91 μV 90 μV 0.14 mV 0.27 mV 0.62 mV 0.68 mV 2.5 mV 12 mV 18 mV 48 mV 0.10 V 0.18 V</p>		B



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
AC VOLTAGE Measurement (continued)	1 V to 5 V 10Hz 10 Hz to 2 kHz 2 kHz to 10 kHz. 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 1 MHz to 2 MHz 2 MHz to 4 MHz 4 MHz to 8 MHz 8 MHz to 10 MHz 5 V to 10 V 10Hz 10 Hz to 2 kHz 2 kHz to 10 kHz. 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 1 MHz to 2 MHz 2 MHz to 4 MHz 4 MHz to 8 MHz 8 MHz to 10 MHz 10 V to 50 V 10Hz 10 Hz to 2 kHz 2 kHz to 10 kHz. 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 50 V to 100 V 10 Hz to 2 kHz 2 kHz to 10 kHz. 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 100 V to 500 V 40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 500 V to 1000 V 40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	0.49 mV 0.48 mV 0.73 mV 1.3 mV 3.3 mV 14 mV 60 mV 0.10 V 0.26 V 0.49 V 0.90 V 0.91 mV 0.90 mV 1.4 mV 2.7 mV 6.3 mV 25 mV 0.12 V 0.18 V 0.48 V 1.0 V 1.8 V 6.0 mV 5.6 mV 6.7 mV 16 mV 39 mV 0.22 V 0.88 V 11 mV 13 mV 28 mV 71 mV 0.43 V 1.4 V 73 mV 0.16 V 0.38 V 0.13 V 0.28 V 0.70 V		B



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
AC CURRENT Generation	10 μA to 50 μA 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 50 μA to 100 μA 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 100 μA to 220 μA 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 0.22 mA to 0.5 mA 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 0.5 mA to 1 mA 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 1 mA to 2.2 mA 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 2.2 mA to 5 mA 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 5 mA to 10 mA 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 mA to 22 mA 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	38 nA 26 nA 71 nA 0.16 μ A 57 nA 30 nA 0.10 μ A 0.24 μ A 0.11 μ A 48 nA 0.19 μ A 0.47 μ A 48 nA 47 nA 0.58 μ A 1.2 μ A 0.49 μ A 0.19 μ A 1.0 μ A 2.4 μ A 1.1 μ A 0.41 μ A 1.9 μ A 4.7 μ A 2.5 μ A 1.0 μ A 7.0 μ A 16 μ A 4.9 μ A 1.9 μ A 10 μ A 24 μ A 11 μ A 4.1 μ A 19 μ A 47 μ A		B



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd

Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
AC CURRENT Generation (continued)	22 mA to 50 mA 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 50 mA to 100 mA 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 100 mA to 220 mA 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 0.22 A to 0.5 A 20 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 0.5 A to 1 A 20 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 1 A to 2.2 A 20 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 2.2 A to 3 A 10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 3 A to 11 A 45 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 5 kHz 11 A to 20.5 A 45 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 5 kHz 45 Hz to 400 Hz 20 A to 30 A	25 μ A 11 μ A 70 μ A 0.16 mA 49 μ A 21 μ A 99 μ A 0.24 mA 0.11 mA 46 μ A 0.19 mA 0.47 mA 0.44 mA 0.50 mA 0.62 mA 0.87 mA 0.99 mA 1.2 mA 1.9 mA 2.2 mA 2.6 mA 6.2 mA 2.1 mA 21 mA 87 mA 8.0 mA 13 mA 0.38 mA 29 mA 36 mA 710 mA 150 mA	Calibration of current clamps using a multi turn coil	B



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
AC CURRENT Generation (continued)	30 A to 110 A <i>45 Hz to 100 Hz</i>	520 mA	Calibration of current clamps using a multi turn coil	B
	110 A to 200 A <i>100 Hz to 400 Hz</i>	960 mA		
	200 A to 550 A <i>45 Hz to 100 Hz</i>	2.5 A		
	550 A to 1025 A <i>45 Hz to 100 Hz</i>	4.8 A		
Measurement	0 μA to 5 μA <i>10 Hz to 55 Hz</i> <i>55 Hz to 10 kHz</i>	13 nA 25 nA		
	5 μA to 10 μA <i>10 Hz to 55 Hz</i> <i>55 Hz to 10 kHz</i>	24 nA 34 nA		
	10 μA to 20 μA <i>10 Hz to 55 Hz</i> <i>55 Hz to 10 kHz</i>	47 nA 52 nA		
	20 μA to 50 μA <i>10 Hz to 2 kHz.</i> <i>2 kHz to 10 kHz</i> <i>10 kHz to 20 kHz</i> <i>20 kHz to 30 kHz</i>	23 nA 35 nA 48 nA 53 nA		
	50 μA to 100 μA <i>10 Hz to 2 kHz.</i> <i>2 kHz to 10 kHz</i> <i>10 kHz to 20 kHz</i> <i>20 kHz to 30 kHz</i>	36 nA 63 nA 88 nA 91 nA		
	100 μA to 200 μA <i>10 Hz to 2 kHz.</i> <i>2 kHz to 10 kHz</i> <i>10 kHz to 30 kHz</i>	67 nA 0.12 μ A 0.17 μ A		
	0.2 mA to 0.5 mA <i>10Hz to 2 kHz.</i> <i>2 kHz to 10 kHz</i> <i>10 kHz to 30 kHz</i>	0.19 μ A 0.32 μ A 0.44 μ A		
	0.5 mA to 1 mA <i>10 Hz to 2 kHz.</i> <i>2 kHz to 10 kHz</i> <i>10 kHz to 30 kHz</i>	0.34 μ A 0.62 μ A 0.86 μ A		



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
AC CURRENT Measurement (continued)	1 mA to 2 mA 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz 2 mA to 5 mA 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz 5 mA to 10 mA 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz 10 mA to 20 mA 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz 20 mA to 50 mA 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz 50 mA to 100 mA 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz 100 mA to 200 mA 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz 0.2 A to 0.5 A 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz 0.5 A to 1 A 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz 1 A to 2 A 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz 2 A to 5 A 10 Hz to 10 kHz	0.66 μ A 1.2 μ A 1.7 μ A 1.9 μ A 3.2 μ A 4.4 μ A 3.4 μ A 6.2 μ A 8.6 μ A 6.6 μ A 12 μ A 17 μ A 19 μ A 31 μ A 45 μ A 34 μ A 61 μ A 86 μ A 66 μ A 0.12 mA 0.17 mA 0.23 mA 0.36 mA 0.47mA 0.38 mA 0.66 mA 0.92 mA 0.74 mA 1.3 mA 1.8 mA 5.0 mA		B



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
AC CURRENT Measurement (continued)	5 A to 10 A <i>10 Hz to 10 kHz</i>	10 mA		B
	10 A to 20 A <i>10 Hz to 10 kHz</i>	19 mA		
	20 A to 30 A <i>10 Hz to 2 kHz</i> <i>2 kHz to 10 kHz</i>	24 mA 32 mA		
FREQUENCY				B
Generation	10 MHz	3.2 in 10^9	Reference	
Measurement	10 Hz to 100 kHz 100 kHz to 1 MHz 1 MHz to 10 MHz 10 MHz to 50 MHz 50 MHz to 600 MHz 600 MHz to 6 GHz	16 mHz 170 mHz 1.7 Hz 17 Hz 4.8 Hz 35 Hz		
Generation	1 Hz to 10 Hz 10 Hz to 100 Hz 100 Hz to 20 MHz	2.3 μ Hz/Hz 1.2 μ Hz/Hz 0.012 μ Hz/Hz	May be reported as events per unit time	
Elapsed time			Mechanical timers / stop watches	
Single event	1 s to 12 hrs	0.05 s		
Revolutions Per Minute Simulation	6 RPM to 600 RPM 600 RPM to 6000 RPM 6000 RPM to 120000 RPM	0.064 RPM 0.065 RPM 0.066 RPM	Optical	
CAPACITANCE				B
Simulated generation	1 kHz 220 pF to 400 pF 400 pF to 1.1 nF 1.1 nF to 3.3 nF 3.3 nF to 11 nF 11 nF to 33 nF 33 nF to 110 nF 110 nF to 330 nF	12 pF 13 pF 22 pF 34 pF 100 pF 320 pF 970 pF		



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
CAPACITANCE (continued) Simulated generation (continued)	100 Hz 0.33 μ F to 1.1 μ F 1.1 μ F to 3.3 μ F 3.3 μ F to 11 μ F 11 μ F to 33 μ F 33 μ F to 110 μ F 110 μ F to 330 μ F 330 μ F to 1.1 mF 1.1 mF to 3.3 mF 3.3 mF to 11 mF 11 mF to 33 mF 33 mF to 110 mF	3.4 nF 10 nF 34 nF 160 nF 600 nF 1.8 μ F 5.8 μ F 18 μ F 58 μ F 290 μ F 1.4 mF		B
Measurement	0 to 0.5 nF 0.5 nF to 1 nF 1 nF to 2 nF 2 nF to 5 nF 5 nF to 10 nF 10 nF to 20 nF 20 nF to 50 nF 50 nF to 100 nF 100 nF to 200 nF 0.2 μ F to 0.5 μ F 0.5 μ F to 1 μ F 1 μ F to 2 μ F 2 μ F to 5 μ F 5 μ F to 10 μ F 10 μ F to 20 μ F 20 μ F to 50 μ F 50 μ F to 100 μ F 100 μ F to 200 μ F 0.2 mF to 0.5 mF 0.5 mF to 1 mF 1 mF to 2 mF 2 mF to 5 mF 5 mF to 10 mF 10 mF to 20 mF 20 mF to 50 mF 50 mF to 100 mF 100 mF to 200 mF	2.0 pF 2.7 pF 5.1 pF 6.1 pF 10 pF 20 pF 35 pF 60 pF 0.12 nF 0.28 nF 0.50 nF 1.1 nF 3.5 nF 5.5 nF 10 nF 45 nF 76 nF 0.15 μ F 0.45 μ F 0.76 μ F 1.5 μ F 5.0 μ F 8.7 μ F 17 μ F 45 μ F 87 μ F 0.17 mF		



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
CALIBRATION OF INSULATION TESTERS				B
Insulation Resistance	10 k Ω to 190 k Ω 200 k Ω 400 k Ω 800 k Ω 1.9 M Ω 2 M Ω 4 M Ω 8 M Ω 19 M Ω 20 M Ω 40 M Ω 80 M Ω 190 M Ω 200 M Ω 400 M Ω 800 M Ω 1.9 G Ω	5.8 k Ω 6.1 k Ω 6.2 k Ω 6.4 k Ω 7.0 k Ω 9.4 k Ω 15 k Ω 28 k Ω 58 k Ω 0.13 M Ω 0.24 M Ω 0.47 M Ω 0.58 M Ω 1.8 M Ω 3.4 M Ω 6.7 M Ω 13 M Ω	Up to 1350 V	B
Insulation test voltage Nominal	0 to 10 V 10 V to 100 V 100 V to 1100 V	0.10 V 0.26 V 2.6 V	The test voltage will normally be measured with a 1 mA load.	B
Continuity Resistance	100 m Ω to 2 Ω 4 Ω 8 Ω 19 Ω 20 Ω 40 Ω 80 Ω 190 Ω 200 Ω 400 Ω 800 Ω 1.9 k Ω 2 k Ω 4 k Ω 8 k Ω 10 k Ω	30 m Ω 33 m Ω 35 m Ω 38 m Ω 55 m Ω 0.12 Ω 0.20 Ω 0.24 Ω 0.74 Ω 1.1 Ω 2.0 Ω 2.4 Ω 8.1 Ω 11 Ω 20 Ω 24 Ω		B



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)		Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.					
Temperature indicators, calibration by electrical simulation				By injection of equivalent DC voltages	B
Type K thermocouple	-270 °C to -200 °C -200 °C to -50 °C -50 °C to 1370 °C	Excluding CJC 1.6 °C 0.10 °C 0.10 °C	Including CJC 2.8 °C 0.17 °C 0.13 °C		
Type J thermocouple	-210 °C to -200 °C -200 °C to 0 °C 0 °C to 1200 °C	0.10 °C 0.10 °C 0.10 °C	0.23 °C 0.21 °C 0.14 °C		
Type E thermocouple	-270 °C to -200 °C -200 °C to 0 °C 0 °C to 1000 °C	0.69 °C 0.10 °C 0.10 °C	0.95 °C 0.13 °C 0.12 °C		
Type T thermocouple	-270 °C to -200 °C -200 °C to 0 °C 0 °C to 400 °C	1.1 °C 0.10 °C 0.10 °C	7.2 °C 0.48 °C 0.22 °C		
Type N thermocouple	-270 °C to -200 °C -200 °C to -100 °C -100 °C to 800 °C 800 °C to 1300 °C	3.7 °C 0.13 °C 0.10 °C 0.10 °C	4.3 °C 0.18 °C 0.13 °C 0.12 °C		
Type R thermocouple	-50 °C to 0 °C 0 °C to 150 °C 150 °C to 400 °C 400 °C to 1768 °C	0.30 °C 0.22 °C 0.15 °C 0.12 °C	0.34 °C 0.25 °C 0.19 °C 0.17 °C		
Type S thermocouple	-50 °C to 0 °C 0 °C to 100 °C 100 °C to 300 °C 300 °C to 1768 °C	0.28 °C 0.21 °C 0.16 °C 0.13 °C	0.31 °C 0.24 °C 0.20 °C 0.17 °C		
Resistance thermometer (Pt 100)	-200 °C to -80 °C -80 °C to 0 °C 0 °C to 100 °C 100 °C to 300 °C 300 °C to 400 °C 400 °C to 630 °C 630 °C to 800 °C	0.058 °C 0.058 °C 0.081 °C 0.10 °C 0.12 °C 0.14 °C 0.27 °C		By injection of equivalent DC resistance	



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)		Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.					
Temperature simulators, calibration by electrical measurement				By measurement of equivalent DC voltages	B
Type K thermocouple	-270 °C to -200 °C	Excluding CJC 1.2 °C	Including CJC 2.6 °C		
	-200 °C to -50 °C	0.10 °C	0.17 °C		
	-50 °C to 1370 °C	0.10 °C	0.13 °C		
Type J thermocouple	-210 °C to -200 °C	0.10 °C	0.23 °C		
	-200 °C to 0 °C	0.10 °C	0.21 °C		
	0 °C to 1200 °C	0.10 °C	0.14 °C		
Type E thermocouple	-270 °C to -200 °C	0.54 °C	0.85 °C		
	-200 °C to 0 °C	0.10 °C	0.13 °C		
	0 °C to 1000 °C	0.10 °C	0.12 °C		
Type T thermocouple	-270 °C to -200 °C	0.86 °C	7.2 °C		
	-200 °C to 0 °C	0.10 °C	0.47 °C		
	0 °C to 400 °C	0.10 °C	0.22 °C		
Type N thermocouple	-270 °C to -200 °C	2.9 °C	3.6 °C		
	-200 °C to -100 °C	0.11 °C	0.16 °C		
	-100 °C to 800 °C	0.10 °C	0.13 °C		
	800 °C to 1300 °C	0.10 °C	0.12 °C		
Type R thermocouple	-50 °C to 0 °C	0.24 °C	0.29 °C		
	0 °C to 150 °C	0.17 °C	0.22 °C		
	150 °C to 400 °C	0.12 °C	0.17 °C		
	400 °C to 1768 °C	0.10 °C	0.15 °C		
Type S thermocouple	-50 °C to 0 °C	0.22 °C	0.26 °C		
	0 °C to 100 °C	0.17 °C	0.21 °C		
	100 °C to 300 °C	0.13 °C	0.17 °C		
	300 °C to 1768 °C	0.11 °C	0.16 °C		
Resistance thermometer (Pt 100)	-200 °C to 0 °C 0 °C to 850 °C	0.0069 °C 0.020 °C		By measurement of equivalent DC resistance	



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd

Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory references unless otherwise stated in the remarks column.				
Calibration of Oscilloscopes				B
Voltage deflection	1 mV to 25 mV 25 mV to 110 mV 110 mV to 2.2 V 2.2 V to 6 V	Q [0.29 %, 47 μ V] Q [0.29 %, 48 μ V] Q [0.29 %, 92 μ V] Q[0.29 %, 0.21 mV]	Into 50 Ω	
	1 mV to 25 mV 25 mV to 110 mV 110 mV to 2.2 V 2.2 V to 11 V 11 V to 130 V	Q [0.12 %, 46 V] Q [0.12 %, 46 μ V] Q [0.12 %, 51 μ V] Q [0.12 %, 0.22 mV] Q [0.12 %, 0.47 mV]	Into 1 M Ω	
Time markers	2 ns to 500 ns 500 ns to 500 μ s 500 μ s to 50 ms 50 ms to 5 s	Q [0.082 %, 0.12 ps] Q [0.082 %, 0.12 ns] Q [0.058 %, 12 ns] Q [0.59 %, 1.2 μ s]		
PRESSURE				B
Pneumatic Pressure (gauge)			Methods consistent with EURAMET CG17. Absolute pressures can be generated within these gauge pressure ranges. This will attract an additional uncertainty of 10 Pa.	
Calibration of pressure indicating instruments and gauges	- 95 kPa to -10 kPa -10 kPa to -3.5 kPa -3.5 kPa to -1.5 kPa 1.5 kPa to 3.5 kPa 3.5 kPa to 10 kPa 10 kPa to 20 kPa 20 kPa to 2.5 MPa 2.5 MPa to 14 MPa	0.0087 % 0.0097 % Q [0.014 %, 0.50 Pa] Q [0.013 %, 0.50 Pa] 0.0078 % 0.0066 % 0.0063 % 0.0079 %	Sensors with an electrical output can be calibrated	
Hydraulic Pressure (gauge)				
Calibration of pressure indicating instruments and gauges	0.6 MPa to 6 MPa 6 MPa to 120 MPa	0.011 % 0.010 %		
END				



0654

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Calibration, Maintenance and Repair Ltd
Issue No: 043 Issue date: 23 August 2024

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$