


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|--|--|--|
|  <p>UKAS TESTING</p> <p>28144</p> <p>Accredited to ISO/IEC 17025:2017</p> | <p>Forth Scientific Ltd</p> <p>Issue No: 001 Issue date: 08 May 2024</p> | |
| | <p>4 Bellman Way Donibristle Industrial Estate Dalgety Bay Dunfermline KY11 9JW</p> | <p>Contact: Clive Ovens Tel: +44 (0) 7894309995 E-Mail: clive@forthscientific.co.uk Website: www.forthscientific.co.uk</p> |
| <p>Testing performed at the above address only</p> | | |

DETAIL OF ACCREDITATION

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used |
|---|--|--|
| <p>WATERS (NON-REGULATORY)</p> <p>Drinking Water, Raw Surface Water, Recreational (Man Made), Process Water (Cooling Towers)</p> | <p><u>Microbiological Tests</u></p> <p>Enumeration:</p> <p>Coliform and <i>E.coli</i> bacteria, confirmed</p> <p>Aerobic Colony Count at: 3 days 22°C and 1 and 2 days, 37°C</p> <p>Enterococci, confirmed</p> <p><i>Pseudomonas aeruginosa</i>, confirmed</p> <p>Legionella spp, presumptive and identification of Legionella spp: <i>Legionella pneumophila</i> sg 1, 2-14 and species</p> | <p>Documented In-house Method</p> <p>LM02 using membrane filtration based on the Microbiology of Drinking Water Part 4 (2016) part B</p> <p>LM01 using Pour Plate based on the Microbiology of Drinking Water Part 7 (2020)</p> <p>LM05 using membrane filtration based on the Microbiology of Drinking Water Part 5 (2012)</p> <p>LM03 using membrane filtration based on the Microbiology of Drinking Water Part 8 (2015)</p> <p>LM04 based on ISO 11731:2017 using filtration with washing, using GVPC, identification using latex agglutination (Matrix A and B); procedure 8,9 and 10 Media A and C</p> |
| <p>END</p> | | |