


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> REFERENCE MATERIALS</p> <p><b>5710</b></p> <p>Accredited to <b>ISO 17034:2016</b></p>	<p><b>EffectTech Limited</b></p> <p><b>Issue No:</b> 005    <b>Issue date:</b> 25 September 2024</p>	
	<p><b>Dove House</b> <b>Dove Fields</b> <b>Uttoxeter</b> <b>Staffordshire</b> <b>ST14 8HU</b></p>	<p><b>Contact: Adam Lomax</b> <b>Tel: +44 (0)1889 569229</b> <b>E-Mail: adam.lomax@effectech.co.uk</b> <b>Website: www.effectech.co.uk</b></p>
<p><b>Reference material production at the above address</b></p>		

### DETAIL OF ACCREDITATION

Matrix / Artefact	Property Value(s) / Identity / Characterisation Range	Characterisation Procedure / Technique	Type* (CRM / RM)	
SYNTHETIC NATURAL GAS MIXTURES	amount fraction	(% mol/mol)	<p><b>In-house method TM001/UT</b></p> <p>Calibration of certified reference materials (CRM) by ISO 6143:2001 using gas chromatography.</p>	CRM
	nitrogen	(0.1 to 22)		
	carbon dioxide	(0.05 to 15)		
	methane	(34 to 100)		
	ethane	(0.1 to 35)		
	propane	(0.05 to 15)		
	iso-butane	(0.01 to 2)		
	n-butane	(0.01 to 2)		
	neo-pentane	(0.002 to 0.35)		
	iso-pentane	(0.005 to 0.35)		
	n-pentane	(0.005 to 0.35)		
	n-hexane	(0.001 to 0.35)		
	2-methylpentane	(0.001 to 0.35)		
	3-methylpentane	(0.001 to 0.35)		
	2,2-dimethylbutane	(0.001 to 0.35)		
	benzene	(0.001 to 0.2)		
	cyclohexane	(0.001 to 0.2)		
	n-heptane	(0.001 to 0.2)		
	toluene	(0.001 to 0.1)		
	methylcyclohexane	(0.001 to 0.1)		
	n-octane	(0.0005 to 0.05)		
n-nonane	(0.0005 to 0.02)			
n-decane	(0.0005 to 0.005)			
helium	(0.005 to 0.2)			
hydrogen	(0.05 to 20)			
oxygen	(0.05 to 3)			



5710

Accredited to  
ISO 17034:2016

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

#### EffecTech Limited

Issue No: 005 Issue date: 25 September 2024

Reference material production performed at the main address only

Matrix / Artefact	Property Value(s) / Identity / Characterisation Range	Characterisation Procedure / Technique	Type* (CRM / RM)
SYNTHETIC NATURAL GAS MIXTURES (cont'd)	amount fraction (% mol/mol)	<b>In-house method TM026/UT</b>	CRM
	oxygen (0.001 to 22.5)	Calibration of certified reference materials (CRM) by ISO 12963:2017 using galvanic fuel cell sensors	
	superior calorific value molar basis (kJ.mol <sup>-1</sup> ) mass basis (MJ.kg <sup>-1</sup> ) volume basis (MJ.m <sup>-3</sup> ) inferior calorific value molar basis (kJ.mol <sup>-1</sup> ) mass basis (MJ.kg <sup>-1</sup> ) volume basis (MJ.m <sup>-3</sup> ) relative density density (kg.m <sup>-3</sup> ) superior Wobbe index (MJ.m <sup>-3</sup> ) inferior Wobbe index (MJ.m <sup>-3</sup> ) molar mass (kg.kmol <sup>-1</sup> ) compression factor	Values calculated by <b>ISO 6976:1995</b> (including amendment No 1, May 1998) on a <i>real</i> or <i>ideal</i> gas basis assuming mixture is dry (free from water)  Combustion properties can be expressed in units of the Joule (J) or in kilowatt hours (kWh)	CRM
gross calorific value molar basis (kJ.mol <sup>-1</sup> ) mass basis (MJ.kg <sup>-1</sup> ) volume basis (MJ.m <sup>-3</sup> ) net calorific value molar basis (kJ.mol <sup>-1</sup> ) mass basis (MJ.kg <sup>-1</sup> ) volume basis (MJ.m <sup>-3</sup> ) relative density density (kg.m <sup>-3</sup> ) gross Wobbe index (MJ.m <sup>-3</sup> ) net Wobbe index (MJ.m <sup>-3</sup> ) molar mass (kg.kmol <sup>-1</sup> ) compression factor	Values calculated by <b>ISO 6976:2016</b> on a <i>real</i> or <i>ideal</i> gas basis assuming mixture is dry (free from water)  Combustion properties can be expressed in units of the Joule (J) or in kilowatt hours (kWh)	CRM	



5710

Accredited to  
ISO 17034:2016

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### EffecTech Limited

Issue No: 005 Issue date: 25 September 2024

Reference material production performed at the main address only

Matrix / Artefact	Property Value(s) / Identity / Characterisation Range	Characterisation Procedure / Technique	Type* (CRM / RM)
SYNTHETIC NATURAL GAS MIXTURES (cont'd)	gross heating value net heating value relative density compressibility factor	Values calculated by methods given in <b>GPA 2172-19</b> (2019) using data tables from <b>GPA 2145-16</b>	CRM
	gross heating value net heating value relative density density compressibility factor	Values calculated by methods given in <b>ASTM D3588-98</b> (2017) using data tables from <b>GPA 2145-16</b>	CRM
SULPHUR GAS MIXTURES	amount fraction (µmol/mol)	<b>In-house method TM002/UT</b>	CRM
	hydrogen sulphide (0.2 to 10)	Calibration of certified reference materials (CRM) using gas chromatography with sulphur chemiluminescence detection (GC-SCD)	CRM
	carbonyl sulphide (0.2 to 10)		
	methanethiol (methyl mercaptan) (0.2 to 10)		
	ethanethiol (ethyl mercaptan) (0.2 to 10)		
	dimethyl sulphide (0.2 to 10)		
	1-propanethiol (n-propyl mercaptan) (0.2 to 10)		
	2-propanethiol (iso-propyl mercaptan) (0.2 to 10)		
	ethyl methyl sulphide (methyl ethyl sulphide) (0.2 to 10)		
	1-butanethiol (n-butyl mercaptan) (0.2 to 10)		
	2-methyl-2-propanethiol (tert-butyl mercaptan) (0.2 to 10)		
	2-methyl-1-propanethiol (iso-butyl mercaptan) (0.2 to 10)		
	1-methyl-1-propanethiol (sec-butyl mercaptan) (0.2 to 10)		
	diethyl sulphide (0.2 to 10)		
n-hexyl mercaptan (0.2 to 10)			
tetrahydrothiophene (THT) (0.2 to 10)			



5710

Accredited to  
ISO 17034:2016

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**EffecTech Limited**

**Issue No: 005 Issue date: 25 September 2024**

Reference material production performed at the main address only

Matrix / Artefact	Property Value(s) / Identity / Characterisation Range	Characterisation Procedure / Technique	Type* (CRM / RM)
BLAST FURNACE GAS MIXTURES (BFG)	amount fraction (% mol/mol)	<b>In-house method TM004/UT</b>  Calibration of certified reference materials (CRM) by ISO 6143:2001 using gas chromatography with thermal conductivity detection (GC-TCD)	CRM
	nitrogen (27 to 54)		
	carbon dioxide (20 to 31)		
	hydrogen (1 to 16)		
	carbon monoxide (20 to 31)		
BASIC OXYGEN STEELMAKING GAS (BOS)	amount fraction (% mol/mol)	<b>In-house method TM004/UT</b>  Calibration of certified reference materials (CRM) by ISO 6143:2001 using gas chromatography with thermal conductivity detection (GC-TCD)	CRM
	nitrogen (11 to 30)		
	carbon dioxide (12 to 20)		
	carbon monoxide (45 to 75)		
	hydrogen (0.4 to 3)		
	oxygen (0.3 to 1.3)		
BINARY EMISSION GAS MIXTURES	amount fraction (% mol/mol)	<b>In-house method TM014</b>  Calibration of certified reference materials (CRM) by ISO 12963:2017 using dynamically generated reference gases in accordance with ISO 6145 Part 7: Thermal Mass Flow Controllers	CRM
	oxygen in nitrogen (0.5 to 25)		
	methane in nitrogen (0.1 to 5)		
	methane in synthetic air (0.1 to 2.5)		
	amount fraction (µmol/mol)		
	carbon monoxide in nitrogen or synthetic air (10 to 1000)		CRM
	nitric oxide in nitrogen (10 to 600)		
	nitrogen dioxide in synthetic air (5 to 500)		
	sulphur dioxide in nitrogen or synthetic air (10 to 1000)		



5710

Accredited to  
ISO 17034:2016

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**EffecTech Limited**

**Issue No: 005 Issue date: 25 September 2024**

Reference material production performed at the main address only

Matrix / Artefact	Property Value(s) / Identity / Characterisation Range	Characterisation Procedure / Technique	Type* (CRM / RM)
BINARY EMISSION GAS MIXTURES (cont'd)	amount fraction                      (% mol/mol)  oxygen in nitrogen                              (0.001 to 22.5)	<b>In-house method TM026/UT</b>  Calibration of certified reference materials (CRM) by ISO 12963:2017 using galvanic fuel cell sensors	CRM
END			

**\* Type**

CRM= Certified Reference Material(s)

RM = Reference Material(s)

*Refer to ISO 17034 for full definitions*