


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

|  |  |  |
|--|--|--|
|  <p><b>UKAS</b><br/>CALIBRATION</p> <p><b>0666</b></p> <p>Accredited to<br/><b>ISO/IEC 17025:2017</b></p> | <p><b>Allied Aerosystems Ltd</b></p> <p><b>Issue No: 058    Issue date: 03 September 2024</b></p>                    |  |
|  | <p><b>Avionic and Metrology Centre</b><br/>Unit G1/2<br/>Treforest Industrial Estate<br/>Pontypridd<br/>CF37 5YL</p> | <p><b>Contact: Mr P Ashurst</b><br/>Tel: +44 (0)1443 849970<br/>Fax: +44 (0)1443 849988<br/>E-Mail: <a href="mailto:phillip.ashurst@allied-aerosystems.com">phillip.ashurst@allied-aerosystems.com</a><br/>Website: <a href="http://www.allied-aerosystems.co.uk">www.allied-aerosystems.co.uk</a></p> |
| <p><b>Calibration performed by the Organisations at the locations specified below</b></p>  |  |  |

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

| Location details   | Activity  | Location code |
|--|---|---------------|
| <p><b>Address</b><br/>Avionic and Metrology Centre<br/>Unit G1/2<br/>Treforest Industrial Estate<br/>Pontypridd<br/>CF37 5YL</p> <p><b>Local contact</b><br/>Mr P Ashurst<br/>Tel: +44 (0)1443 849970<br/>Fax: +44 (0)1443 849988<br/>E-Mail:<br/><a href="mailto:phillip.ashurst@allied-aerosystems.com">phillip.ashurst@allied-aerosystems.com</a></p> | <p>Electrical, DC and LF<br/>Electrical, RF<br/>Rotational speed<br/>Time</p>   | Lab 1         |
| <p><b>Address</b><br/>Units 1-6, Moy Road<br/>Industrial Estate<br/>Taffs Well<br/>Cardiff<br/>CF15 7QR</p> <p><b>Local contact</b><br/>All correspondence and deliveries are to be made to the above address</p>  | <p>Dimensional<br/>Torque<br/>Mass<br/>Temperature<br/>Relative Humidity<br/>Pressure<br/>Surface plates and tables<br/>Accelerometry</p> | Lab 2         |

#### Site activities performed away from the locations listed above:

| Location details   | Activity   | Location code |
|--|--|---------------|
| <p>Customers' Premises</p> <p>The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.</p> | <p>Non-automatic weighing instruments<br/>Surface plates and tables<br/>Temperature<br/>Pressure</p> | Site          |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)

| Measured Quantity<br>Instrument or Gauge | Range  | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )  | Remarks  | Location<br>Code |
|--|--|---|--|------------------|
| ELECTRICAL CALIBRATION                   |  |   |  |                  |
| DC VOLTAGE                               | 0 mV to 220 mV<br>220 mV to 2.2 V<br>2.2 V to 11 V<br>11 V to 22 V<br>22 V to 220 V<br>220 V to 1100 V   | 8.4 $\mu\text{V/V} + 0.60 \mu\text{V}$<br>7.0 $\mu\text{V/V} + 1.0 \mu\text{V}$<br>7.0 $\mu\text{V/V} + 3.6 \mu\text{V}$<br>7.0 $\mu\text{V/V} + 6.6 \mu\text{V}$<br>8.0 $\mu\text{V/V} + 80 \mu\text{V}$<br>9.0 $\mu\text{V/V} + 560 \mu\text{V}$    | Using multifunction calibrator.  | Lab 1            |
|  | 0 mV to 200 mV<br>200 mV to 2 V<br>2 V to 20 V<br>20 V to 200 V<br>200 V to 1 kV   | 5.2 $\mu\text{V/V} + 1.2 \mu\text{V}$<br>3.3 $\mu\text{V/V} + 0.59 \mu\text{V}$<br>3.2 $\mu\text{V/V} + 5.1 \mu\text{V}$<br>4.7 $\mu\text{V/V} + 51 \mu\text{V}$<br>15 $\mu\text{V/V} + 0.29 \text{mV}$   | Using digital multimeter   |                  |
| DC CURRENT                               | 0 $\mu\text{A}$ to 220 $\mu\text{A}$<br>220 $\mu\text{A}$ to 2.2 mA<br>2.2 mA to 22 mA<br>22 mA to 220 mA<br>220 mA to 2.2 A<br>2.2 A to 11 A<br>11 A to 20 A<br>20 A to 100 A | 50 $\mu\text{A/A} + 8.0 \text{nA}$<br>50 $\mu\text{A/A} + 8.0 \text{nA}$<br>50 $\mu\text{A/A} + 100 \text{nA}$<br>60 $\mu\text{A/A} + 800 \text{nA}$<br>81 $\mu\text{A/A} + 25 \mu\text{A}$<br>0.080 % + 0.39 mA<br>0.10 % + 4.6 mA<br>0.10 % + 23 mA | Using multifunction calibrator.  | Lab 1            |
|  | 11 A to 16.5 A<br>16.5 A to 150 A<br>150 A to 1025 A   | 0.30 % + 2.4 mA<br>0.30 % + 18 mA<br>0.30 % + 61 mA   | Simulated current using a multi turn coil, for the calibration of clamp-on ammeters and similar devices. |                  |
|  | 0 $\mu\text{A}$ to 200 $\mu\text{A}$<br>200 $\mu\text{A}$ to 2 mA<br>2 mA to 20 mA<br>20 mA to 200 mA<br>200 mA to 2 A<br>2.0 A to 20 A<br>20 A to 100 A                       | 10 $\mu\text{A/A} + 0.48 \text{nA}$<br>10 $\mu\text{A/A} + 4.7 \text{nA}$<br>10 $\mu\text{A/A} + 47 \text{nA}$<br>39 $\mu\text{A/A} + 0.93 \mu\text{A}$<br>200 $\mu\text{A/A} + 16 \mu\text{A}$<br>440 $\mu\text{A/A} + 51 \mu\text{A}$<br>0.11 %     | Using digital multimeter.  |                  |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )  | Remarks  | Location<br>Code |
|--|---|---|--|------------------|
| ELECTRICAL CALIBRATION<br>(continued)    |   |   |  |                  |
| DC RESISTANCE                            |   |   |  | Lab 1            |
| Specific values                          | 0 $\Omega$<br>1 m $\Omega$<br>10 m $\Omega$<br>100 m $\Omega$<br>1.0 $\Omega$<br>1.9 $\Omega$<br>10 $\Omega$<br>19 $\Omega$<br>100 $\Omega$<br>190 $\Omega$<br>1.0 k $\Omega$<br>1.9 k $\Omega$<br>10 k $\Omega$<br>19 k $\Omega$<br>100 k $\Omega$<br>190 k $\Omega$<br>1.0 M $\Omega$<br>1.9 M $\Omega$<br>10 M $\Omega$<br>19 M $\Omega$<br>100 M $\Omega$ | 59 $\mu\Omega$<br>100 $\mu\Omega/\Omega$<br>100 $\mu\Omega/\Omega$<br>100 $\mu\Omega/\Omega$<br>110 $\mu\Omega/\Omega$<br>98 $\mu\Omega/\Omega$<br>31 $\mu\Omega/\Omega$<br>37 $\mu\Omega/\Omega$<br>18 $\mu\Omega/\Omega$<br>18 $\mu\Omega/\Omega$<br>14 $\mu\Omega/\Omega$<br>14 $\mu\Omega/\Omega$<br>12 $\mu\Omega/\Omega$<br>13 $\mu\Omega/\Omega$<br>15 $\mu\Omega/\Omega$<br>16 $\mu\Omega/\Omega$<br>25 $\mu\Omega/\Omega$<br>30 $\mu\Omega/\Omega$<br>54 $\mu\Omega/\Omega$<br>68 $\mu\Omega/\Omega$<br>110 $\mu\Omega/\Omega$ | Known values of DC<br>resistance for application to<br>measuring instruments in a 2-<br>wire or 4-wire configuration,<br>as appropriate. |                  |
|  | 0 $\Omega$ to 2 $\Omega$<br>2 $\Omega$ to 20 $\Omega$<br>20 $\Omega$ to 200 $\Omega$<br>200 $\Omega$ to 2 k $\Omega$<br>2 k $\Omega$ to 20 k $\Omega$<br>20 k $\Omega$ to 200 k $\Omega$<br>200 k $\Omega$ to 2 M $\Omega$<br>2 M $\Omega$ to 20 M $\Omega$<br>20 M $\Omega$ to 200 M $\Omega$<br>200 M $\Omega$ to 2 G $\Omega$                              | 12 $\mu\Omega/\Omega + 5.0 \mu\Omega$<br>8.5 $\mu\Omega/\Omega + 21 \mu\Omega$<br>8.5 $\mu\Omega/\Omega + 70 \mu\Omega$<br>8.4 $\mu\Omega/\Omega + 0.62 \text{ m}\Omega$<br>8.4 $\mu\Omega/\Omega + 6.2 \text{ m}\Omega$<br>8.6 $\mu\Omega/\Omega + 62 \text{ m}\Omega$<br>9.9 $\mu\Omega/\Omega + 1.4 \Omega$<br>16 $\mu\Omega/\Omega + 120 \Omega$<br>68 $\mu\Omega/\Omega + 670 \Omega$<br>0.060 % + 1.2 M $\Omega$  | Using digital multimeter; for<br>the calibration of resistors and<br>resistance boxes.   |                  |
| AC VOLTAGE                               | 10 Hz to 20 Hz<br>1 mV to 2.2 mV<br>2.2 mV to 22 mV<br>22 mV to 220 mV<br>220 mV to 2.2 V<br>2.2 V to 22 V<br>22 V to 220 V<br><br>20 Hz to 40 Hz<br>1 mV to 2.2 mV<br>2.2 mV to 22 mV<br>22 mV to 220 mV<br>220 mV to 2.2 V<br>2.2 V to 22 V<br>22 V to 220 V  | 0.071 % + 4.6 $\mu\text{V}$<br>0.057 % + 5.0 $\mu\text{V}$<br>0.057 % + 13 $\mu\text{V}$<br>0.051 % + 80 $\mu\text{V}$<br>0.051 % + 810 $\mu\text{V}$<br>0.051 % + 8.1 mV<br><br>0.050 % + 4.5 $\mu\text{V}$<br>0.023 % + 5.0 $\mu\text{V}$<br>0.022 % + 8.0 $\mu\text{V}$<br>0.017 % + 25 $\mu\text{V}$<br>0.017 % + 250 $\mu\text{V}$<br>0.017 % + 2.6 mV   | Using multifunction calibrator.  | Lab 1            |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge                        | Range  | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )   | Remarks                         | Location<br>Code |
|---|--|--|---------------------------------|------------------|
| ELECTRICAL CALIBRATION<br>(continued)<br>AC VOLTAGE (continued) | <p><i>40 Hz to 20 kHz</i><br/>1 mV to 2.2 mV<br/>2.2 mV to 22 mV<br/>22 mV to 220 mV<br/>220 mV to 2.2 V<br/>2.2 V to 22 V<br/>22 V to 220 V</p> <p><i>50 Hz to 1 kHz</i><br/>220 V to 1100 V</p> <p><i>20 kHz to 50 kHz</i><br/>1 mV to 2.2 mV<br/>2.2 mV to 22 mV<br/>22 mV to 220 mV<br/>220 mV to 2.2 V<br/>2.2 V to 22 V<br/>22 V to 220 V</p> <p><i>50 kHz to 100 kHz</i><br/>1 mV to 2.2 mV<br/>2.2 mV to 22 mV<br/>22 mV to 220 mV<br/>220 mV to 2.2 V<br/>2.2 V to 22 V<br/>22 V to 220 V</p> <p><i>100 kHz to 300 kHz</i><br/>1 mV to 2.2 mV<br/>2.2 mV to 22 mV<br/>22 mV to 220 mV<br/>220 mV to 2.2 V<br/>2.2 V to 22 V</p> <p><i>300 kHz to 500 kHz</i><br/>1 mV to 2.2 mV<br/>2.2 mV to 22 mV<br/>22 mV to 220 mV<br/>220 mV to 2.2 V<br/>2.2 V to 22 V</p> <p><i>500 kHz to 1 MHz</i><br/>1 mV to 2.2 mV<br/>2.2 mV to 22 mV<br/>22 mV to 220 mV<br/>220 mV to 2.2 V<br/>2.2 V to 22 V</p> | <p>0.046 % + 4.5 <math>\mu</math>V<br/>0.014 % + 5.0 <math>\mu</math>V<br/>0.0090 % + 1.0 <math>\mu</math>V<br/>0.0090 % + 7.0 <math>\mu</math>V<br/>0.0090 % + 62 <math>\mu</math>V<br/>0.0090 % + 1.0 mV</p> <p>0.010 % + 5.0 mV</p> <p>0.058 % + 6.0 <math>\mu</math>V<br/>0.039 % + 7.0 <math>\mu</math>V<br/>0.033 % + 8.0 <math>\mu</math>V<br/>0.013 % + 17 <math>\mu</math>V<br/>0.013 % + 170 <math>\mu</math>V<br/>0.023 % + 3.5 mV</p> <p>0.10 % + 7.0 <math>\mu</math>V<br/>0.090 % + 7.0 <math>\mu</math>V<br/>0.090 % + 25 <math>\mu</math>V<br/>0.025 % + 70 <math>\mu</math>V<br/>0.025 % + 350 <math>\mu</math>V<br/>0.050 % + 8.0 mV</p> <p>0.12 % + 13 <math>\mu</math>V<br/>0.11 % + 12 <math>\mu</math>V<br/>0.11 % + 25 <math>\mu</math>V<br/>0.050 % + 140 <math>\mu</math>V<br/>0.050 % + 1.5 mV</p> <p>0.18 % + 25 <math>\mu</math>V<br/>0.17 % + 25 <math>\mu</math>V<br/>0.17 % + 35 <math>\mu</math>V<br/>0.11 % + 360 <math>\mu</math>V<br/>0.13 % + 4.3 mV</p> <p>0.35 % + 25 <math>\mu</math>V<br/>0.34 % + 25 <math>\mu</math>V<br/>0.34 % + 80 <math>\mu</math>V<br/>0.22 % + 850 <math>\mu</math>V<br/>0.27 % + 8.5 mV</p> | Using multifunction calibrator. | Lab 1            |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )  | Remarks                   | Location<br>Code |
|--|---|---|---------------------------|------------------|
| AC VOLTAGE (continued)                   | <p><i>10 Hz to 20 Hz</i><br/>1.3 <math>\mu</math>V to 2.2 mV<br/>2.2 mV to 7 mV<br/>7 mV to 22 mV<br/>22 mV to 70 mV<br/>70 mV to 220 mV<br/>220 mV to 700 mV<br/>700 mV to 2.2 V<br/>2.2 V to 7 V<br/>7 V to 22 V<br/>22 V to 70 V<br/>70 V to 220 V</p> <p><i>20 Hz to 40 Hz</i><br/>2 <math>\mu</math>V to 2.2 mV<br/>2.2 mV to 7 mV<br/>7 mV to 22 mV<br/>22 mV to 70 mV<br/>70 mV to 220 mV<br/>220 mV to 700 mV<br/>700 mV to 2.2 V<br/>2.2 V to 7 V<br/>7 V to 22 V<br/>22 V to 70 V<br/>70 V to 220 V</p> <p><i>40 Hz to 20 kHz</i><br/>1.3 <math>\mu</math>V to 2.2 mV<br/>2.2 mV to 7 mV<br/>7 mV to 22 mV<br/>22 mV to 70 mV<br/>70 mV to 220 mV<br/>220 mV to 700 mV<br/>700 mV to 2.2 V<br/>2.2 V to 7 V<br/>7 V to 22 V<br/>22 V to 70 V<br/>70 V to 220 V</p> <p><i>100 Hz to 20 kHz</i><br/>220 V to 700 V<br/>700 V to 1100</p> <p><i>20 kHz to 50 kHz</i><br/>1.8 <math>\mu</math>V to 2.2 mV<br/>2.2 mV to 7 mV<br/>7 mV to 22 mV<br/>22 mV to 70 mV<br/>70 mV to 220 mV<br/>220 mV to 700 mV<br/>700 mV to 2.2 V<br/>2.2 V to 7 V<br/>7 V to 22 V<br/>22 V to 70 V<br/>70 V to 220 V<br/>220 V to 700 V</p> | <p>0.010 % 1.3 <math>\mu</math>V<br/>70 <math>\mu</math>V/V + 1.8 <math>\mu</math>V<br/>60 <math>\mu</math>V/V + 2.1 <math>\mu</math>V<br/>50 <math>\mu</math>V/V + 1.5 <math>\mu</math>V<br/>50 <math>\mu</math>V/V + 2.4 <math>\mu</math>V<br/>50 <math>\mu</math>V/V + 11 <math>\mu</math>V<br/>50 <math>\mu</math>V/V + 23 <math>\mu</math>V<br/>50 <math>\mu</math>V/V + 0.11 mV<br/>50 <math>\mu</math>V/V + 0.3 mV<br/>50 <math>\mu</math>V/V + 1.7 mV<br/>50 <math>\mu</math>V/V + 3.7 mV</p> <p>50 <math>\mu</math>V/V + 2.0 <math>\mu</math>V<br/>50 <math>\mu</math>V/V + 2.2 <math>\mu</math>V<br/>50 <math>\mu</math>V/V + 2.1 <math>\mu</math>V<br/>30 <math>\mu</math>V/V + 2.2 <math>\mu</math>V<br/>30 <math>\mu</math>V/V + 3 <math>\mu</math>V<br/>30 <math>\mu</math>V/V + 10.5 <math>\mu</math>V<br/>30 <math>\mu</math>V/V + 14.6 <math>\mu</math>V<br/>30 <math>\mu</math>V/V + 30.3 <math>\mu</math>V<br/>30 <math>\mu</math>V/V + 0.28 mV<br/>30 <math>\mu</math>V/V + 0.55 mV<br/>30 <math>\mu</math>V/V + 2.0 mV</p> <p>40 <math>\mu</math>V/V + 1.3 <math>\mu</math>V<br/>30 <math>\mu</math>V/V + 1.5 <math>\mu</math>V<br/>30 <math>\mu</math>V/V + 1.8 <math>\mu</math>V<br/>20 <math>\mu</math>V/V + 1.7 <math>\mu</math>V<br/>20 <math>\mu</math>V/V + 2.5 <math>\mu</math>V<br/>20 <math>\mu</math>V/V + 13 <math>\mu</math>V<br/>20 <math>\mu</math>V/V + 12 <math>\mu</math>V<br/>20 <math>\mu</math>V/V + 44 <math>\mu</math>V<br/>20 <math>\mu</math>V/V + 81 <math>\mu</math>V<br/>20 <math>\mu</math>V/V + 0.13 mV<br/>40 <math>\mu</math>V/V + 0.90 mV</p> <p>20 <math>\mu</math>V/V + 12 mV<br/>20 <math>\mu</math>V/V + 20 mV</p> <p>50 <math>\mu</math>V/V + 1.8 <math>\mu</math>V<br/>40 <math>\mu</math>V/V + 1.9 <math>\mu</math>V<br/>40 <math>\mu</math>V/V + 2.1 <math>\mu</math>V<br/>30 <math>\mu</math>V/V + 3.2 <math>\mu</math>V<br/>20 <math>\mu</math>V/V + 4 <math>\mu</math>V<br/>20 <math>\mu</math>V/V + 7 <math>\mu</math>V<br/>20 <math>\mu</math>V/V + 15 <math>\mu</math>V<br/>20 <math>\mu</math>V/V + 25 <math>\mu</math>V<br/>20 <math>\mu</math>V/V + 41 <math>\mu</math>V<br/>30 <math>\mu</math>V/V + 0.61 mV<br/>30 <math>\mu</math>V/V + 1.9 mV<br/>90 <math>\mu</math>V/V + 15 mV</p> | Using digital multimeter. | Lab 1            |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge                        | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )   | Remarks   | Location<br>Code |
|---|---|--|---|------------------|
| ELECTRICAL CALIBRATION<br>(continued)<br>AC VOLTAGE (continued) | <p><i>50 kHz to 100 kHz</i><br/>2.2 <math>\mu</math>V to 2.2 mV<br/>2.2 mV to 7 mV<br/>7 mV to 22 mV<br/>22 mV to 70 mV<br/>70 mV to 220 mV<br/>220 mV to 700 mV<br/>700 mV to 2.2 V<br/>2.2 V to 7 V<br/>7 V to 22 V<br/>22 V to 70 V<br/>70 V to 220 V<br/>220 V to 700 V</p> <p><i>100 kHz to 300 kHz</i><br/>5.7 <math>\mu</math>V to 2.2 mV<br/>2.2 mV to 7 mV<br/>7 mV to 22 mV<br/>22 mV to 70 mV<br/>70 mV to 220 mV<br/>220 mV to 700 mV<br/>700 mV to 2.2 V<br/>2.2 V to 7 V<br/>7 V to 22 V<br/>22 V to 70 V</p> <p><i>300 kHz to 500 kHz</i><br/>6 <math>\mu</math>V to 2.2 mV<br/>2.2 mV to 7 mV<br/>7 mV to 22 mV<br/>22 mV to 70 mV<br/>70 mV to 220 mV<br/>220 mV to 700 mV<br/>700 mV to 2.2 V<br/>2.2 V to 7 V<br/>7 V to 22 V</p> <p><i>500 kHz to 1 MHz</i><br/>7 <math>\mu</math>V to 2.2 mV<br/>2.2 mV to 7 mV<br/>7 mV to 22 mV<br/>22 mV to 70 mV<br/>70 mV to 220 mV<br/>220 mV to 700 mV<br/>700 mV to 2.2 V<br/>2.2 V to 7 V<br/>7 V to 22 V</p> | <p>70 <math>\mu</math>V/V + 2.2 <math>\mu</math>V<br/>50 <math>\mu</math>V/V + 2.2 <math>\mu</math>V<br/>70 <math>\mu</math>V/V + 2.5 <math>\mu</math>V<br/>70 <math>\mu</math>V/V + 4.4 <math>\mu</math>V<br/>50 <math>\mu</math>V/V + 9.3 <math>\mu</math>V<br/>50 <math>\mu</math>V/V + 3.7 <math>\mu</math>V<br/>50 <math>\mu</math>V/V + 25 <math>\mu</math>V<br/>60 <math>\mu</math>V/V + 31 <math>\mu</math>V<br/>60 <math>\mu</math>V/V + 21 <math>\mu</math>V<br/>60 <math>\mu</math>V/V + 4.0 mV<br/>60 <math>\mu</math>V/V + 9.1 mV<br/>0.040 % + 17 mV</p> <p>0.017 % + 5.7 <math>\mu</math>V<br/>0.010 % + 5.8 <math>\mu</math>V<br/>0.0090 % + 5.8 <math>\mu</math>V<br/>0.010 % + 6.9 <math>\mu</math>V<br/>0.0090 % + 11 <math>\mu</math>V<br/>0.0090 % + 9 <math>\mu</math>V<br/>0.0080 % + 24 <math>\mu</math>V<br/>0.012 % + 67 <math>\mu</math>V<br/>0.012 % + 21 <math>\mu</math>V<br/>0.012 % + 7.0 mV</p> <p>0.051 % + 6.0 <math>\mu</math>V<br/>0.032 % + 7.0 <math>\mu</math>V<br/>0.027 % + 9.0 <math>\mu</math>V<br/>0.023 % + 22 <math>\mu</math>V<br/>0.022 % + 9.0 <math>\mu</math>V<br/>0.022 % + 34 <math>\mu</math>V<br/>0.019 % + 17 <math>\mu</math>V<br/>0.030 % + 0.18 mV<br/>0.030 % + 0.20 mV</p> <p>0.19 % + 7.0 <math>\mu</math>V<br/>0.14 % + 7.0 <math>\mu</math>V<br/>0.11 % + 9.0 <math>\mu</math>V<br/>0.11 % + 21 <math>\mu</math>V<br/>0.080 % + 7.0 <math>\mu</math>V<br/>0.080 % + 33 <math>\mu</math>V<br/>0.070 % + 21 <math>\mu</math>V<br/>0.10 % + 0.34 mV<br/>0.10 % + 0.26 mV</p> | <p>Using digital multimeter.</p> <p>Using digital multimeter.</p> | Lab 1            |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge                | Range  | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )   | Remarks  | Location<br>Code |
|---|--|--|--|------------------|
| ELECTRICAL CALIBRATION<br>(continued)<br><br>AC CURRENT |  |  |  | Lab 1            |
|   | <i>45 Hz to 1 kHz</i><br>22 $\mu$ A to 220 $\mu$ A<br>220 $\mu$ A to 2.2 mA<br>2.2 mA to 22 mA<br>22 mA to 220 mA<br>220 mA to 2.2 A         | 0.021 % + 16 nA<br>0.018 % + 35 nA<br>0.014 % + 370 nA<br>0.014 % + 3.6 $\mu$ A<br>0.065 % + 42 $\mu$ A                            | Using multifunction calibrator.  |                  |
|   | <i>45 Hz to 100 Hz</i><br>2.2 A to 11 A<br>11 A to 20 A<br>20 A to 100 A   | 0.070 % + 2.4 mA<br>0.14 % + 5.8 mA<br>0.27 % + 120 mA   |  |                  |
|   | <i>100 Hz to 1 kHz</i><br>2.2 A to 11 A<br>11 A to 20 A  | 0.070 % + 2.4 mA<br>0.17 % + 24 mA   |  |                  |
|   | <i>45 Hz to 65 Hz</i><br>11 A to 16.5 A<br>16.5 A to 150 A<br>150 A to 1025 A  | 0.40 % + 3.5 mA<br>0.40 % + 29 mA<br>0.40 % + 110 mA   | Simulated current using a multi turn coil, for the calibration of clamp-on ammeters and similar devices. |                  |
|   | <i>65 Hz to 440 Hz</i><br>11 A to 16.5 A<br>16.5 A to 150 A<br>150 A to 1025 A   | 1.0 % + 4.0 mA<br>1.0 % + 32 mA<br>1.0 % + 120 mA  | Simulated current using a multi turn coil, for the calibration of clamp-on ammeters and similar devices. |                  |
|   | <i>1 kHz to 5 kHz</i><br>22 $\mu$ A to 220 $\mu$ A<br>220 $\mu$ A to 2.2 mA<br>2.2 mA to 22 mA<br>22 mA to 220 mA<br>220 mA to 2.2 A         | 0.064 % + 40 nA<br>0.062 % + 400 nA<br>0.060 % + 4.0 $\mu$ A<br>0.060 % + 40 $\mu$ A<br>0.075 % + 80 $\mu$ A                       |  |                  |
|   | <i>5 kHz to 10 kHz</i><br>22 $\mu$ A to 220 $\mu$ A<br>220 $\mu$ A to 2.2 mA<br>2.2 mA to 22 mA<br>22 mA to 220 mA<br>220 mA to 2.2 A        | 0.17 % + 120 nA<br>0.17 % + 800 nA<br>0.16 % + 8.0 $\mu$ A<br>0.16 % + 80 $\mu$ A<br>0.85 % + 160 $\mu$ A                          |  |                  |
|   | <i>50 Hz to 1 kHz</i><br>1 $\mu$ A to 200 $\mu$ A<br>200 $\mu$ A to 2 mA<br>2 mA to 20 mA<br>20 mA to 200 mA<br>200 mA to 2 A<br>2 A to 20 A | 0.034 % + 26 nA<br>0.033 % + 0.26 $\mu$ A<br>0.033 % + 2.6 $\mu$ A<br>0.034 % + 26 $\mu$ A<br>0.078 % + 0.26 mA<br>0.11 % + 2.6 mA | Using digital multimeter.  |                  |
|   | <i>50 Hz to 60 Hz</i><br>20 A to 100 A   | 0.17 %   |  |                  |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )   | Remarks   | Location<br>Code |
|--|---|--|---|------------------|
| ELECTRICAL CALIBRATION<br>(continued)    |   |  |   |                  |
| AC CURRENT (continued)                   | <i>1 kHz to 5 kHz</i><br>1 $\mu$ A to 200 $\mu$ A<br>200 $\mu$ A to 2 mA<br>2 mA to 20 mA<br>20 mA to 200 mA<br>200 mA to 2 A<br>2 A to 20 A<br><br><i>5 kHz to 10 kHz</i><br>1 $\mu$ A to 200 $\mu$ A<br>200 $\mu$ A to 2 mA<br>2 mA to 20 mA<br>20 mA to 200 mA<br>200 mA to 2 A<br>2 A to 20 A | 0.036 % + 26 nA<br>0.033 % + 0.26 $\mu$ A<br>0.033 % + 6.4 $\mu$ A<br>0.034 % + 26 $\mu$ A<br>0.090 % + 0.27 mA<br>0.33 % + 2.6 mA<br><br>0.052 % + 26 nA<br>0.034 % + 0.26 $\mu$ A<br>0.033 % + 6.4 $\mu$ A<br>0.034 % + 26 $\mu$ A<br>0.091 % + 0.27 mA<br>0.38 % + 2.0 mA | Using digital multimeter.   | Lab 1            |
| DC POWER                                 |   |  |   | Lab 1            |
| 0.1 mW to 100 kW                         | 33 mV to 1020 V<br>3.3 mA to 9 mA<br>9 mA to 33 mA<br>33 mA to 90 mA<br>90 mA to 330 mA<br>330 mA to 900 mA<br>900 mA to 2.2 A<br>2.2 A to 4.5 A<br>4.5 A to 11 A<br>11 A to 20.5 A<br>20.5 A to 100 A  | 0.040 %<br>0.030 %<br>0.040 %<br>0.030 %<br>0.080 %<br>0.060 %<br>0.12 %<br>0.090 %<br>0.13 %<br>0.13 %  | For voltages in the<br>range 33 mV to<br>1020 V.                              |                  |
| 0.1 W to 1 MW                            | 10 A to 16.5 A<br>16.5 A to 110 A<br>110 A to 150 A<br>150 A to 1000 A  | 0.25 %<br>0.27 %<br>0.28 %<br>0.31 %   | For voltages in the<br>range 33 mV to<br>1020 V; clamp meter<br>calibrations. |                  |





0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ ) | Remarks   | Location<br>Code |   |       |
|--|---|--|---|------------------|---|-------|
| ELECTRICAL CALIBRATION<br>(continued)    |   |  |   |                  |   |       |
| AC POWER                                 |   |  | By phantom load technique at<br>unity power factor. | Lab 1            |   |       |
| 0.1 mW to 100 kW<br>45 Hz to 65 Hz       | 33 mV to 330 mV<br>3.3 mA to 9 mA<br>9 mA to 33 mA<br>33 mA to 90 mA<br>90 mA to 330 mA   | 0.40 %<br>0.25 %<br>0.35 %<br>0.25 %               |   |                  |   |       |
|  | 330 mA to 900 mA<br>900 mA to 2.2 A<br>2.2 A to 4.5 A<br>4.5 A to 11 A                    | 0.35 %<br>0.25 %<br>0.35 %<br>0.30 %               |   |                  |   |       |
|  | 11 A to 20.5 A<br>20.5 A to 100 A   | 0.22 %<br>0.90 %                                   |   |                  |   |       |
|  | 330 mV to 1020 V<br>3.3 mA to 9 mA<br>9 mA to 33 mA<br>33 mA to 90 mA<br>90 mA to 330 mA  | 0.25 %<br>0.15 %<br>0.25 %<br>0.15 %               |   |                  |   |       |
|  | 330 mA to 900 mA<br>900 mA to 2.2 A<br>2.2 A to 4.5 A<br>4.5 A to 11 A                    | 0.25 %<br>0.15 %<br>0.30 %<br>0.30 %               |   |                  |   |       |
|  | 11 A to 20.5 A<br>20.5 A to 100 A   | 0.22 %<br>0.90 %                                   |   |                  |   |       |
| 0.1 W to 1 MW                            | 33 mV to 1020 V<br>10 A to 16.5 A<br>16.5 A to 110 A<br>110 A to 150 A<br>150 A to 1000 A | 0.32 %<br>0.34 %<br>0.34 %<br>0.43 %               |   |                  | Clamp meter calibrations.   |       |
| PHASE ANGLE                              |   |  |   |                  |   | Lab 1 |
| Voltage:Current                          | 0° to 360°<br>10 Hz to 65 Hz<br>65 Hz to 400 Hz   | 0.18°<br>1.0°                                      |   |                  | For applied currents in the<br>range 3.3 mA to 20.5 A and<br>applied voltages in the range<br>33 mV to 1020 V |       |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range  | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )                         | Remarks   | Location<br>Code |
|--|--|--|---|------------------|
| ELECTRICAL CALIBRATION<br>(continued)    |  |  |   |                  |
| CAPACITANCE                              |  |  |   | Lab 1            |
| Calibration of capacitance<br>meters     |  |  | Using multifunction calibrator.   |                  |
|  | 10 Hz to 10 kHz<br>220 pF to 1.0999 nF   | 0.39 % + 9.8 pF  | <p>The frequency ranges shown are for the test frequencies for conventional LCR meters or for the sample rate of capacitance meters that operate using the principle <math>C = I\Delta t / \Delta V</math>.</p> <p>The frequency ranges shown are for the sample rate of capacitance meters that operate using the principle <math>C = I\Delta t / \Delta V</math>.</p> |                  |
|  | 10 Hz to 3 kHz<br>1.1 nF to 3.2999 nF  | 0.39 % + 9.8 pF  |   |                  |
|  | 10 Hz to 1 kHz<br>3.3 nF to 10.999 nF<br>11 nF to 32.999 nF<br>33 nF to 109.99 nF<br>110 nF to 329.99 nF | 0.20 % + 13 pF<br>0.20 % + 0.10 nF<br>0.20 % + 0.13 nF<br>0.20 % + 0.42 nF |   |                  |
|  | 10 Hz to 600 Hz<br>0.33 $\mu$ F to 1.0999 $\mu$ F  | 0.20 % + 2.8 nF  |   |                  |
|  | 10 Hz to 300 Hz<br>1.1 $\mu$ F to 3.2999 $\mu$ F   | 0.20 % + 3.9 nF  |   |                  |
|  | 10 Hz to 150 Hz<br>3.3 $\mu$ F to 10.999 $\mu$ F   | 0.20 % + 13 nF   |   |                  |
|  | 10 Hz to 120 Hz<br>11 $\mu$ F to 32.999 $\mu$ F  | 0.31 % + 52 nF   |   |                  |
|  | 10 Hz to 80 Hz<br>33 $\mu$ F to 109.99 $\mu$ F   | 0.35 % + 0.18 $\mu$ F  |   |                  |
|  | DC to 50 Hz<br>110 $\mu$ F to 329.99 $\mu$ F   | 0.35 % + 0.65 $\mu$ F  |   |                  |
|  | DC to 20 Hz<br>0.33 mF to 1.0999 mF  | 0.35 % + 1.2 $\mu$ F   |   |                  |
|  | DC to 6 Hz<br>1.1 mF to 3.2999 mF  | 0.35 % + 6.5 $\mu$ F   |   |                  |
|  | DC to 2 Hz<br>3.3 mF to 10.999 mF  | 0.34 % + 12 $\mu$ F  |   |                  |
|  | DC to 0.6 Hz<br>11 mF to 32.999 mF   | 0.58 % + 68 $\mu$ F  |   |                  |
|  | DC to 0.2 Hz<br>33 mF to 110 mF  | 0.85 % + 0.13 mF   |   |                  |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge   | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )   | Remarks  | Location<br>Code |
|--|---|--|--|------------------|
| <b>Calibration of 17<sup>th</sup>/18<sup>th</sup><br/>Edition Test Equipment</b> |   |  |  | Lab 1            |
| <b>INSULATION TESTERS</b>  |   |  |  |                  |
| Continuity   | 0 $\Omega$ to 100 m $\Omega$<br>100 m $\Omega$ to 500 m $\Omega$<br>0.5 $\Omega$ to 1 $\Omega$<br>1 $\Omega$ to 5 $\Omega$<br>5 $\Omega$ to 200 $\Omega$<br>200 $\Omega$ to 10 k $\Omega$   | 5.0 m $\Omega$<br>1.1 % + 12 m $\Omega$<br>0.53 % + 12 m $\Omega$<br>0.46 % + 12 m $\Omega$<br>0.25 % + 12 m $\Omega$<br>0.25 %  | Applied resistances of known value.  |                  |
| High Resistance  | 10 k $\Omega$ to 1 M $\Omega$<br>1 M $\Omega$ to 10 M $\Omega$<br>10 M $\Omega$ to 1 G $\Omega$<br>1 G $\Omega$ to 10 G $\Omega$  | 0.24 %<br>0.35 %<br>0.6 %<br>1.3 %   | Applied resistances of known value. Uncertainty increases by 0.10 % per 200 V above 500 V. |                  |
| Test Voltage Measurement   | 0 V to 2 kV DC<br>10 k $\Omega$ to 1 M $\Omega$ load<br>1 M $\Omega$ to 10 G $\Omega$ load  | 1.0 % + 2.0 V<br>1.0 % + 5.0 V   | By voltage measurement.  |                  |
| <b>LINE / LOOP IMPEDANCE TESTERS</b>   |   |  |  |                  |
| Resistance (nominal values)  | 25 m $\Omega$<br>50 m $\Omega$<br>100 m $\Omega$<br>330 m $\Omega$<br>500 m $\Omega$<br>1 $\Omega$<br>1.8 $\Omega$<br>5 $\Omega$<br>10 $\Omega$<br>18 $\Omega$<br>50 $\Omega$<br>100 $\Omega$<br>180 $\Omega$<br>500 $\Omega$<br>1 k $\Omega$<br>1.8 k $\Omega$ | 6.0 m $\Omega$<br>6.0 m $\Omega$<br>6.0 m $\Omega$<br>9.0 m $\Omega$<br>10 m $\Omega$<br>12 m $\Omega$<br>22 m $\Omega$<br>36 m $\Omega$<br>0.070 $\Omega$<br>0.12 $\Omega$<br>0.35 $\Omega$<br>0.59 $\Omega$<br>1.2 $\Omega$<br>3.0 $\Omega$<br>5.9 $\Omega$<br>12 $\Omega$ | At 50 Hz nominal   |                  |
| Current  | 0.05 A<br>0.1 A<br>0.25 A<br>0.5 A<br>0.8 A<br>1.5 A<br>2.0 A<br>3.2 A<br>6 A<br>8 A<br>10 A<br>14 A<br>25 A<br>28 A<br>30 A  | 1.5 % + 2.0 mA<br>1.5 % + 3.0 mA<br>1.5 % + 5.0 mA<br>1.5 % + 10 mA<br>1.5 % + 20 mA<br>1.5 % + 30 mA<br>1.5 % + 50 mA<br>1.5 % + 70 mA<br>1.5 % + 100 mA<br>1.5 % + 150 mA<br>1.5 % + 0.20 A<br>1.5 % + 0.30 A<br>1.5 % + 0.35 A<br>1.5 % + 0.50 A<br>1.5 % + 0.70 A        | 50 Hz nominal  |                  |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )  | Remarks                          | Location<br>Code |
|--|---|---|----------------------------------|------------------|
| <b>RCD TESTERS</b>                       |   |   |                                  |                  |
| Timing                                   | 10 ms to 5 s  | 0.024 % + 0.32 ms   |                                  |                  |
| Current                                  |   |   |                                  |                  |
| 0.5 x I and 1 x I Mode                   | 3 mA to 30 mA<br>30 mA to 300 mA<br>300 mA to 3000 mA   | 1.3 % + 1.3 $\mu$ A<br>1.3 % + 14 $\mu$ A<br>1.3 % + 0.13 mA  |                                  |                  |
| 1.4 x I and 2 x I Mode                   | 3 mA to 30 mA<br>30 mA to 300 mA<br>300 mA to 1500 mA   | 2.6 % + 1.3 $\mu$ A<br>2.6 % + 14 $\mu$ A<br>2.6 % + 0.13 mA  |                                  |                  |
| 5 x I Mode                               | 3 mA to 30 mA<br>30 mA to 300 mA<br>300 mA to 600 mA  | 6.4 % + 1.3 $\mu$ A<br>6.4 % + 14 $\mu$ A<br>6.4 % + 0.13 mA  |                                  |                  |
| <b>LEAKAGE TESTERS</b>                   |   |   |                                  |                  |
| Current                                  |   |   |                                  |                  |
| Passive/Differential/Substitute Modes    | 0.1 mA to 30 mA   | 0.30 % + 2.0 $\mu$ A (ac + dc) rms  |                                  |                  |
| Active Mode                              | 0.1 mA to 30 mA   | 0.30 % + 1.0 $\mu$ A (ac + dc) rms  |                                  |                  |
| Touch Voltage                            | 250 V Range   | 5.0 % + 3.0 V   |                                  |                  |
| <b>PORTABLE APPLIANCE TESTERS</b>        |   |   |                                  | Lab 1            |
| Earth Bond Resistance.                   | 25 m $\Omega$<br>50 m $\Omega$<br>100 m $\Omega$<br>330 m $\Omega$<br>500 m $\Omega$<br>1 $\Omega$<br>1.8 $\Omega$<br>5 $\Omega$<br>10 $\Omega$<br>18 $\Omega$<br>50 $\Omega$<br>100 $\Omega$<br>180 $\Omega$<br>500 $\Omega$<br>1 k $\Omega$<br>1.8 k $\Omega$ | 6.0 m $\Omega$<br>6.0 m $\Omega$<br>6.0 m $\Omega$<br>9.0 m $\Omega$<br>10 m $\Omega$<br>12 m $\Omega$<br>22 m $\Omega$<br>36 m $\Omega$<br>70 m $\Omega$<br>0.12 $\Omega$<br>0.35 $\Omega$<br>0.59 $\Omega$<br>1.2 $\Omega$<br>3.0 $\Omega$<br>5.9 $\Omega$<br>12 $\Omega$ | Nominal values at 50 Hz to 60 Hz |                  |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge  | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )  | Remarks  | Location<br>Code |
|---|---|---|--|------------------|
| PORTABLE APPLIANCE<br>TESTERS (continued) |   |   |  | Lab 1            |
| Earth Bond Current<br>At 50 Hz to 60 Hz   | 0.05 A<br>0.1 A<br>0.25 A<br>0.5 A<br>0.8 A<br>1.5 A<br>2.0 A<br>3.2 A<br>6 A<br>8 A<br>10 A<br>14 A<br>25 A<br>28 A<br>30 A                              | 1.5 % + 2.0 mA<br>1.5 % + 3.0 mA<br>1.5 % + 5.0 mA<br>1.5 % + 10 mA<br>1.5 % + 20 mA<br>1.5 % + 30 mA<br>1.5 % + 50 mA<br>1.5 % + 70 mA<br>1.5 % + 100 mA<br>1.5 % + 150 mA<br>1.5 % + 0.20 A<br>1.5 % + 0.30 A<br>1.5 % + 0.35 A<br>1.5 % + 0.50 A<br>1.5 % + 0.70 A |  |                  |
| Insulation Resistance                     | 10 k $\Omega$ to 1 M $\Omega$<br>1 M $\Omega$ to 10 M $\Omega$<br>10 M $\Omega$ to 1 G $\Omega$<br>1 G $\Omega$ to 10 G $\Omega$                          | 0.24 %<br>0.35 %<br>0.60 %<br>1.3 %   | Applied resistances of known value. Uncertainty increases by 0.10 % per 200 V above 500 V. |                  |
| Test Voltage                              | 0 V to 2 kV DC<br>10 k $\Omega$ to 1 M $\Omega$ load<br>1 M $\Omega$ to 10 G $\Omega$ load  | 1.0 % + 2.0 V<br>1.0 % + 5.0 V  | By voltage measurement.  |                  |
| Continuity                                | 100 m $\Omega$ to 500 m $\Omega$<br>0.5 $\Omega$ to 1 $\Omega$<br>1 $\Omega$ to 5 $\Omega$<br>5 $\Omega$ to 200 $\Omega$<br>200 $\Omega$ to 10 k $\Omega$ | 1.1 % + 12 m $\Omega$<br>0.53 % + 12 m $\Omega$<br>0.46 % + 12 m $\Omega$<br>0.25 % + 12 m $\Omega$<br>0.25 %   |  |                  |
| HIPOT/HV TESTERS                          |   |   |  |                  |
| HVDC Voltage                              | 0.1 kV to 10 kV<br>10 kV to 20 kV<br>20 kV to 35 kV<br>35 kV to 40 kV   | 0.35 % + 7.0 V<br>2.4 % + 46 V<br>1.3 % + 78 V<br>2.4 % + 97 V  |  |                  |
| HVAC Peak Voltage                         | 50 Hz / 60 Hz<br>100 V to 10 kV<br>10 kV to 40 kV   | 0.60 % + 19 V<br>5.8 % + 100 V  |  |                  |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge                        | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )   | Remarks  | Location<br>Code |
|---|---|--|--|------------------|
| HIPOT/HV TESTERS<br>(continued)                                 |   |  |  | Lab 1            |
| DC Leakage Current  | 0 A to 300 $\mu$ A<br>0.3 mA to 3 mA<br>3 mA to 30 mA<br>30 mA to 300 mA  | 0.37 % + 0.26 $\mu$ A<br>0.24 % + 1.9 $\mu$ A<br>0.24 % + 19 $\mu$ A<br>0.24 % + 0.19 mA   |  |                  |
| AC Leakage Current  | 20 Hz to 400 Hz<br>0 A to 300 $\mu$ A<br>0.3 mA to 3 mA<br>3 mA to 30 mA<br>30 mA to 300 mA   | 0.37 % + 0.26 $\mu$ A<br>0.24 % + 1.9 $\mu$ A<br>0.24 % + 19 $\mu$ A<br>0.24 % + 0.19 mA   |  |                  |
| FREQUENCY   | 1 Hz to 10 Hz<br>10 Hz to 100 Hz<br>100 Hz to 1 kHz<br>1kHz to 10kHz<br>10 kHz to 100 kHz<br>100 kHz to 10 MHz<br>10 MHz to 100 MHz<br>100 MHz to 1 GHz<br>1 GHz to 10 GHz<br>10 GHz to 20 GHz              | 2.0 in $10^6$<br>1.0 in $10^6$<br>1.0 in $10^7$<br>1.0 in $10^8$<br>1.0 in $10^9$<br>1.0 in $10^9$<br>1.6 in $10^{10}$<br>19 in $10^{12}$<br>15 in $10^{12}$<br>2.8 in $10^{12}$ | Results and uncertainties may be quoted as average periodic time for repetitive events.          |                  |
| TEMPERATURE SIMULATION  |   |  |  | Lab 1            |
| Electrical calibration of temperature indicators and simulators |   |  |  |                  |
| Resistance thermometers (Pt 100)                                | -200 $^{\circ}$ C to +400 $^{\circ}$ C<br>400 $^{\circ}$ C to 800 $^{\circ}$ C  | 0.070 $^{\circ}$ C<br>0.070 $^{\circ}$ C   | By resistance simulation. Other RTD types can be calibrated but may attract larger uncertainties |                  |
| Thermocouples   |   |  |  |                  |
| Type J  | -250 $^{\circ}$ C to -100 $^{\circ}$ C<br>-100 $^{\circ}$ C to -30 $^{\circ}$ C<br>-30 $^{\circ}$ C to +190 $^{\circ}$ C<br>190 $^{\circ}$ C to 1200 $^{\circ}$ C   | 0.14 $^{\circ}$ C<br>0.080 $^{\circ}$ C<br>0.074 $^{\circ}$ C<br>0.070 $^{\circ}$ C  | By millivolt simulation, including cold junction compensation.                                   |                  |
| Type K  | -200 $^{\circ}$ C to -100 $^{\circ}$ C<br>-100 $^{\circ}$ C to -25 $^{\circ}$ C<br>-25 $^{\circ}$ C to +250 $^{\circ}$ C<br>250 $^{\circ}$ C to 1000 $^{\circ}$ C<br>1000 $^{\circ}$ C to 1372 $^{\circ}$ C | 0.41 $^{\circ}$ C<br>0.37 $^{\circ}$ C<br>0.36 $^{\circ}$ C<br>0.36 $^{\circ}$ C<br>0.36 $^{\circ}$ C  |  |                  |
| Type R  | -50 $^{\circ}$ C to +50 $^{\circ}$ C<br>50 $^{\circ}$ C to 270 $^{\circ}$ C<br>270 $^{\circ}$ C to 1768 $^{\circ}$ C  | 0.50 $^{\circ}$ C<br>0.36 $^{\circ}$ C<br>0.25 $^{\circ}$ C  |  |                  |
| Type T  | -200 $^{\circ}$ C to -150 $^{\circ}$ C<br>-150 $^{\circ}$ C to 0 $^{\circ}$ C<br>0 $^{\circ}$ C to 120 $^{\circ}$ C<br>120 $^{\circ}$ C to 400 $^{\circ}$ C   | 0.16 $^{\circ}$ C<br>0.12 $^{\circ}$ C<br>0.082 $^{\circ}$ C<br>0.074 $^{\circ}$ C   |  |                  |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )  | Remarks   | Location<br>Code |
|--|---|---|---|------------------|
| TIME INTERVAL<br>Elapsed Time            | 1 s to 24 hrs   | 0.22 s<br><br>0.0010 s  | Mechanically triggered. The quoted uncertainty is particularly dependent on the resolution of the timer being calibrated.<br><br>Electronically triggered   | Lab 1            |
| ROTATIONAL SPEED<br>Tachometers          | 60 rpm to 1 000 rpm<br>1 000 rpm to 10 000 rpm<br>10 000 rpm to 120 000 rpm   | 0.12 rpm<br>0.13 rpm to 0.65 rpm<br>0.65 rpm to 1.7 rpm   | Optical or electrically triggered devices.<br>Mechanical tachometers will attract a larger uncertainty due to resolution.   | Lab 1            |
| VOLTAGE REFLECTION<br>COEFFICIENT        | 0 to 1<br>10 MHz to 8 GHz<br>8 GHz to 18 GHz  | 0.030<br>0.070  | Using R & S ZNB-20<br>The uncertainties are quoted for 50 Ω systems for devices fitted only with Type N Connectors. Results and uncertainties may also be quoted in units of Return Loss (dB) or VSWR .   | Lab 1            |
| RF ATTENUATION                           | 10 MHz to 4 GHz<br>0 dB to 10 dB<br>10 dB to 40 dB<br>40 dB to 50 dB<br>50 dB to 70 dB<br>70 dB to 90 dB<br><br>4 GHz to 12 GHz<br>0 dB to 10 dB<br>10 dB to 20 dB<br>20 dB to 30 dB<br>30 dB to 40 dB<br>40 dB to 50 dB<br><br>12 GHz to 18 GHz<br>0 dB to 10 dB<br>10 dB to 20 dB<br>20 dB to 30 dB<br>30 dB to 40 dB<br>40 dB to 50 dB | 0.036 dB<br>0.040 dB<br>0.050 dB<br>0.080 dB<br>0.090 dB<br><br>0.043 dB<br>0.040 dB<br>0.050 dB<br>0.040 dB<br>0.060 dB<br><br>0.16 dB<br>0.16 dB<br>0.15 dB<br>0.16 dB<br>0.24 dB | Using R & S ZNB-20<br><br>The uncertainties are quoted for 50 Ω systems for devices fitted only with Type N Connectors which present a near match to the 50 Ω measurement system.<br>Measurement of devices presenting a significant mismatch will receive much larger uncertainties. | Lab 1            |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range  | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ ) | Remarks   | Location<br>Code |
|--|--|--|---|------------------|
| RF POWER Reference                       | 1 mW Reference 50 MHz  | 0.0063 mW  |   | Lab 1            |
| RF POWER                                 |  |  | Note: These uncertainties are for type N systems with a source VSWR of 1.20:1 or less. Measurement of devices with a higher mismatch will receive higher measurement uncertainties. | Lab 1            |
|  | 100 kHz to 300 kHz<br>+20 dBm to +10 dBm<br>+10 dBm to -25 dBm | 0.21 dB<br>0.16 dB                                 | Using 438A Power Meter and 8482A Sensor.  |                  |
|  | 300 kHz to 1 MHz<br>+20 dBm to +10 dBm<br>+10 dBm to -25 dBm   | 0.16 dB<br>0.10 dB                                 |   |                  |
|  | 1 MHz to 100 MHz<br>+20 dBm to +10 dBm<br>+10 dBm to -25 dBm   | 0.15 dB<br>0.090 dB                                | Using 438A Power Meter and 8482A Sensor   |                  |
|  | 100 MHz to 3 GHz<br>+20 dBm to +10 dBm<br>+10 dBm to -25 dBm   | 0.16 dB<br>0.090 dB                                |   |                  |
|  | 3 GHz to 4.2 GHz<br>+20 dBm to +10 dBm<br>+10 dBm to -25 dBm   | 0.17 dB<br>0.11 dB                                 |   |                  |
|  | 10 MHz to 30 MHz<br>+20 dBm to +10 dBm<br>+10 dBm to -25 dBm   | 0.16 dB<br>0.10 dB                                 | Using 438A Power Meter and 8481A Sensor   |                  |
|  | 30 MHz to 3.9 GHz<br>+20 dBm to +10 dBm<br>+10 dBm to -25 dBm  | 0.15 dB<br>0.080 dB                                |   |                  |
|  | 4 GHz to 18 GHz<br>+20 dBm to +10 dBm<br>+10 dBm to -25 dBm    | 0.16 dB<br>0.10 dB                                 |   |                  |
|  | 10 MHz to 30 MHz<br>-20 dBm to -30 dBm<br>-30 dBm to -60 dBm   | 0.14 dB<br>0.19 dB                                 | Using 438A Power Meter and 8484A Sensor   |                  |
|  | 30 MHz to 3.9 GHz<br>-20 dBm to -30 dBm<br>-30 dBm to -60 dBm  | 0.13 dB<br>0.12 dB                                 |   |                  |





0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )   | Remarks  | Location<br>Code   |   |
|--|---|--|--|--|---|
| RF POWER (continued)                     | 4 GHz to 12 GHz<br>-20 dBm to -30 dBm<br>-30 dBm to -60 dBm   | 0.14 dB<br>0.14 dB   | Using ML2437A Power Meter<br>and MA 2481B Sensor | Lab 1  |   |
|  | 12 GHz to 18 GHz<br>-20 dBm to -30 dBm<br>-30 dBm to -60 dBm  | 0.15 dB<br>0.15 dB   |  |  |   |
|  | 10 MHz to 50 MHz<br>+20 dBm to -60 dBm  | 0.13 dB  |  |  |   |
|  | 50 MHz to 3 GHz<br>+20 dBm to -60 dBm   | 0.10 dB  |  |  |   |
|  | 3 GHz to 6 GHz<br>+20 dBm to -60 dBm  | 0.11 dB  |  |  |   |
|  | 10 MHz to 6.6 GHz<br>0 dBm to -50 dBm<br>-50 dBm to -70 dBm<br>-70 dBm to -100 dBm<br>-100 dBm to -130 dBm  | 0.088 dB<br>0.091 dB<br>0.096 dB<br>0.10 dB  |  | N5531S Measuring Receiver<br>& Sensor Module   | Lab 1   |
|  | 6.6 GHz to 13.2 GHz<br>0 dBm to -40 dBm<br>-40 dBm to -60 dBm<br>-60 dBm to -90 dBm<br>-90 dBm to -120 dBm  | 0.13 dB<br>0.13 dB<br>0.13 dB<br>0.14 dB   |  |  |   |
|  | 13.2 GHz to 18 GHz<br>0 dBm to -30 dBm<br>-30 dBm to -50 dBm<br>-50 dBm to -90 dBm<br>-90 dBm to -120 dBm   | 0.13 dB<br>0.13 dB<br>0.13 dB<br>0.14 dB   |  |  |   |
|  | 2.5 MHz to 1.3 GHz<br>0 dBm to -30 dBm<br>-30 dBm to -80 dBm<br>-80 dBm to -120 dBm<br>-120 dBm to -130 dBm | 0.14 dB<br>0.13 dB<br>0.16 dB<br>0.23 dB   |  | Tuned RF Level using 8902A<br>measuring receiver   |   |
|  | RF PEAK POWER   | +20 dBm to -40 dBm<br>500 MHz to 800 MHz<br>800 MHz to 1 GHz<br>1 GHz to 1.6 GHz<br>1.6 GHz to 2.5 GHz<br>2.5 GHz to 3.2 GHz<br>3.2 GHz to 3.6 GHz<br>3.6 GHz to 4.2 GHz<br>4.2 GHz to 4.4 GHz<br>4.4 GHz to 4.5 GHz<br>4.5 GHz to 4.6 GHz<br>4.6 GHz to 4.8 GHz<br>4.8 GHz to 5.2 GHz<br>5.2 GHz to 5.4 GHz<br>5.4 GHz to 5.5 GHz<br>5.5 GHz to 5.8 GHz<br>5.8 GHz to 6 GHz |  | 0.10 dB<br>0.11 dB<br>0.12 dB<br>0.13 dB<br>0.12 dB<br>0.11 dB<br>0.10 dB<br>0.11 dB<br>0.14 dB<br>0.15 dB<br>0.14 dB<br>0.13 dB<br>0.14 dB<br>0.13 dB<br>0.12 dB<br>0.11 dB | For the calibration of pulsed<br>RF sources and generators<br>using 4541 Peak Power Meter<br>& 57006 Sensor |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range  | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ ) | Remarks  | Location<br>Code |
|--|--|--|--|------------------|
| Bandwidth at -3dB point                  | 9 kHz to 18 GHz<br>-3 dB from set point                | 0.34 MHz   | Filter shape using 438A<br>Power Meter & 8481A Sensor  | Lab 1            |
| AMPLITUDE MODULATION                     |  |  |  | Lab 1            |
| System 1                                 | Modulation rate 50 Hz to 10 kHz<br>5 % AM to 40 % AM   | 2.0 % of reading<br>+ 0.010 % AM                   | Using 8902A measuring receiver<br>$f_c$ 10 MHz to 1.3 GHz  |                  |
| System 2                                 | 40 % AM to 99 % AM                                     | 2.0 % of reading<br>+ 0.10 % AM                    | $f_c$ 10 MHz to 1.3 GHz  |                  |
|  | 5 % AM to 99 % AM<br>Modulation rate 50 Hz to 10 kHz   | 0.75 % of reading                                  | Using N5531S measuring receiver system<br>$f_c$ 100 kHz to 10 MHz  |                  |
|  | 20 % AM to 99 % AM<br>Modulation rate 50 Hz to 100 kHz | 0.50 % of reading                                  | $f_c$ 10 MHz to 3 GHz  |                  |
|  | 5 % AM to 20 % AM<br>Modulation rate 50 Hz to 100 kHz  | 2.5 % of reading                                   | $f_c$ 10 MHz to 3 GHz  |                  |
|  | 20 % AM to 99 % AM<br>Modulation rate 50 Hz to 100 kHz | 1.5 % of reading                                   | $f_c$ 3 GHz to 18 GHz  |                  |
| FREQUENCY MODULATION                     | 5 % AM to 20 % AM<br>Modulation rate 50 Hz to 100 kHz  | 4.5 % of reading                                   | $f_c$ 3 GHz to 18 GHz  |                  |
| System 1                                 | 0 kHz to 4 kHz<br>Modulation rate 20 Hz to 50 Hz       | 2.0 % + 1.0 Hz                                     | $f_c$ 10 MHz to 1.3 GHz, using 8902A measuring receiver. Not all combinations of carrier deviation and modulation frequencies may be realisable. | Lab 1            |
|  | Modulation rate 50 Hz to 100 kHz                       | 1.0 % + 1.0 Hz                                     |  |                  |
|  | Modulation rate 100 kHz to 200 kHz                     | 5.0 % + 1.0 Hz                                     |  |                  |
|  | 4 kHz to 40 kHz<br>Modulation rate 20 Hz to 50 Hz      | 2.0 % + 10 Hz                                      |  |                  |
|  | Modulation rate 50 Hz to 100 kHz                       | 1.0 % + 10 Hz                                      |  |                  |
|  | Modulation rate 100 kHz to 200 kHz                     | 5.0 % + 10 Hz                                      |  |                  |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge     | Range  | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )   | Remarks   | Location<br>Code |
|--|--|--|---|------------------|
| FREQUENCY MODULATION<br>System 1 (continued) | 40 kHz to 400 kHz<br><i>Modulation rate 50 Hz to 100 kHz</i>   | 1.0 % + 100 Hz   |   | Lab 1            |
| System 2                                     | 200 Hz to 400 kHz<br><i>Modulation rate 20 Hz to 200 kHz</i>   | 5.0 % + 100 Hz<br><br>1.0 % of reading   | Using N5531S measuring receiver system. Not all combinations of carrier deviation and modulation frequencies may be realisable. |                  |
| SPECTRUM ANALYSIS                            | +30 dBm to -100 dBm<br><i>3 GHz to 3 GHz</i><br><i>3 GHz to 6.6 GHz</i><br><i>6.6 GHz to 13.2 GHz</i><br><i>13.2 GHz to 22 GHz</i> | 0.25 dB<br>0.75 dB<br>1.2 dB<br>1.1 dB   | Using N5531S measuring receiver system.   | Lab 1            |
| DISTORTION FACTOR                            | 0 dB to -80 dB   | 1.0 dB   | Includes in-band noise and spurs. All spectral components must lie within the range 20 Hz to 250 kHz.                           | Lab 1            |
| DIMENSIONAL MEASUREMENTS                     |  |  | All linear calibrations may be given in inch units.   | Lab 2            |
| MEASURING INSTRUMENTS AND MACHINES           |  |  |   |                  |
| Micrometers                                  |  |  |   |                  |
| External                                     | BS 870:2008<br>0 mm to 600 mm  | Heads 2.0 $\mu$ m between any two points.<br>Zero set<br>1.0 + (8.0 x length in m) $\mu$ m       |   |                  |
| Internal                                     | BS 959:2008 and above<br>0 mm to 1000 mm   | Heads 2.0 $\mu$ m between any two points.<br>Extension rods<br>1.0 + (8.0 x length in m) $\mu$ m |   |                  |
| Depth  | BS 6468:2008<br>0 mm to 300 mm   | Heads 2.0 $\mu$ m between any two points.<br>Extension rods<br>1.0 + (8.0 x length in m) $\mu$ m |   |                  |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge                             | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )                               | Remarks  | Location<br>Code |
|--|---|--|--|------------------|
| Micrometer heads   | As BS 1734:1951<br>0 to 100   | 1.6 $\mu$ m  |  | Lab 2            |
| Micrometer setting rods/<br>Length gauges, flat &<br>spherical ended | BS 870:2008<br>25 mm to 600 mm  | 1.0 + (5.0 x length in m) $\mu$ m  |  |                  |
| Three point bore   | 2 mm to 178 mm  | Overall performance 4.0 $\mu$ m  | Using ring gauges of a<br>known size   |                  |
| Vernier Gauges   |   |  |  |                  |
| Caliper  | BS 887:2008 and above<br>0 mm to 1500 mm  | Overall performance<br>10 + (30 x length in m) $\mu$ m                           |  |                  |
| Depth  | As BS 6365: 2008<br>0 to 600  |  |  |                  |
| Dial gauges and dial test<br>indicators                              | As BS 907:2008 and<br>BS 2795:1981<br>0 mm to 25 mm   | 2.0 $\mu$ m  |  |                  |
| <u>External and Internal Dial<br/>Callipers</u>                      | <u>0-150mm</u>  | <u>3.5 <math>\mu</math>m</u>   | <u>In house procedure<br/>PRO580</u>   |                  |
| Simple height gauges   | BS EN ISO 13225:2012<br>0 mm to 300mm<br>300 mm to 450mm<br>450 mm to 600mm<br>600 mm to 1000mm | Length measurement error (E)<br>5 $\mu$ m<br>6 $\mu$ m<br>7 $\mu$ m<br>8 $\mu$ m |  |                  |
| Spirit levels  | As BS 958:1968<br>5 seconds to 60 minutes of<br>arc nominal sensitivity                         | 1.5 seconds of arc   |  |                  |
| ACCELEROMETRY  |   |  |  | Lab 2            |
| Portable Accelerometer<br>(brake meters)                             | Horizontal acceleration<br>0 % to 100 %   | 1.0 % of horizontal<br>acceleration  | NOTE Mechanical brake<br>meters only<br><br>Horizontal acceleration<br>expressed as a percentage<br>(%) of the local free fall<br>acceleration due to gravity.<br><br>Brake meters calibrated<br>using an inclinable test rig. |                  |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge                      | Range  | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )   | Remarks   | Location<br>Code |
|---|--|--|---|------------------|
| <b>LENGTH</b>   |  |  |   | Lab 2            |
| Plain plug gauges (parallel)<br>Cylindrical setting standards | 1 mm to 50 mm diameter<br>50 mm to 150 mm diameter   | 1.0 $\mu$ m<br>1.5 $\mu$ m   | All linear calibrations may be given in inch units.<br>By comparison with reference standards   |                  |
| Precision pins  | 0.05 mm to 10 mm diameter  | 1.0 $\mu$ m  |   |                  |
| Plain ring gauges (parallel)<br>and setting standards         | 3 mm to 50 mm diameter<br>50 mm to 100 mm<br>100 mm to 150 mm diameter<br>150 mm to 200 mm diameter                                | 1.7 $\mu$ m<br>1.8 $\mu$ m<br>2.2 $\mu$ m<br>2.3 $\mu$ m   | By comparison with reference standards  |                  |
| Thread measuring cylinders                                    | As BS3777:1964 and<br>BS5590:1978 and specials<br>0.1 to 5.0 diameter  | 0.80 $\mu$ m   |   |                  |
| Plain gap gauges (parallel)                                   | 1 to 100mm   | 3.0 $\mu$ m  | Using gauge blocks  |                  |
| Gauge blocks  |  | Class (see note)   |   |                  |
| Inch<br>(Steel)   | As BS 4311-1:2007<br>0.01 inch to 0.4 inch<br>0.4 inch to 1 inch<br>2 inch<br>3 inch<br>4 inch                                     | <u>C</u><br>3.4 $\mu$ inches<br>4.1 $\mu$ inches<br>5.7 $\mu$ inches<br>7.6 $\mu$ inches<br>9.6 $\mu$ inches | Note: <b>Class C</b> uncertainties apply to the measurement of length of steel gauges by comparison with grade K standards of length of a similar material. Class C uncertainties apply to new and used grade 0, 1 & 2 gauges to BS 4311-1:2007 and BS EN ISO 3650:1999 |                  |
| Millimetre<br>(Steel)   | As BS EN ISO 3650:1999<br>0.25 mm to 10 mm<br>10 mm to 25 mm<br>30 mm, 40 mm, 50 mm<br>60 mm, 70 mm, 75 mm<br>80 mm, 90 mm, 100 mm | <u>C</u><br>0.10 $\mu$ m<br>0.11 $\mu$ m<br>0.13 $\mu$ m<br>0.17 $\mu$ m<br>0.20 $\mu$ m                     |   |                  |
| Feeler Gauges   | As BS 957:2008<br>0.02 mm to 1.00 mm   | 1.5 $\mu$ m  |   |                  |
| Paint thickness setting foils                                 | 0.01 mm to 8 mm  | 2.0 $\mu$ m  | By comparison with reference standards and a single axis measuring machine.   |                  |
| Parallels   | As BS 906:1972<br>5 to 50 x 100 x 400  | 2.0 $\mu$ m to 5.0 $\mu$ m   |   |                  |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge          | Range  | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )              | Remarks   | Location<br>Code |
|---|--|---|---|------------------|
| Squares<br>Blade type                             | 50 mm to 450 mm  | 5.5 $\mu$ m   | The uncertainty quoted is for the departure from flatness, i.e. the distance separating the two parallel planes which just enclose the surface under consideration. | Lab 2 and Site   |
| Right angle and box angle plates                  | As BS 5535:1978<br>50 mm to 300 mm                     | Squareness:<br>3.0 + (1.0 per 100 mm)<br>Parallelism: 0.0030 mm |   |                  |
| Vee blocks  | As BS 3731:1987<br>20 mm to 100 mm                     | 5.0 $\mu$ m   |   |                  |
| FORM  |  |   |   |                  |
| Surface plates and tables<br>Granite<br>Cast iron | BS 817:2008<br>160 mm x 100 mm to<br>2500 mm x 1600 mm | 1.5 $\mu$ m +<br>(0.80 x diagonal in m) $\mu$ m<br>See Note 1   | The uncertainty quoted is for the departure from flatness, i.e. the distance separating the two parallel planes which just enclose the surface under consideration. |                  |
| ANCILLARY<br>MEASUREMENTS                         |  |   |   |                  |
| Flatness  |  | 0.2   | Ancillary measurements made for completeness of calibration. Best CMC's are dependent on methodology and range.   | Lab 2            |
| Parallelism                                       |  | 0.8   |   |                  |
| Squareness  |  | 2.5   |   |                  |
| Straightness                                      |  | 1.4 + (0.8 x length in m)                                       |   |                  |
| MASS<br>See notes 1 and 2                         | Nominal Value (g)                                      | (mg)  | Note 1: Calibrations can be given in other units as required.   |                  |
|   | 50 000   | 50  | Note 2: Intermediate values can be calibrated to an uncertainty interpolated from the next higher and lower values in the table.                                    |                  |
|   | 20 000   | 20  |   |                  |
|   | 10 000   | 10  |   |                  |
|   | 5 000  | 5.0   |   |                  |
|   | 2 000  | 2.0   |   |                  |
|   | 1 000  | 1.0   |   |                  |
|   | 500  | 0.50  |   |                  |
|   | 200  | 0.20  |   |                  |
|   | 100  | 0.10  |   |                  |
|   | 50   | 0.060   |   |                  |
|   | 20   | 0.050   |   |                  |
|   | 10   | 0.040   |   |                  |
|   | 5  | 0.030   |   |                  |
|   | 2  | 0.024   |   |                  |
|   | 1  | 0.020   |   |                  |
|   | 0.5  | 0.016   |   |                  |
|   | 0.2  | 0.012   |   |                  |
|   | 0.1  | 0.010   |   |                  |
|   | 0.05   | 0.0080  |   |                  |
|   | 0.02   | 0.0060  |   |                  |
|   | 0.01   | 0.0050  |   |                  |
|   | 0.005 to 0.001   | 0.0040  | Note 3: The calibration is based on Borda's method of substitution  |                  |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge   | Range  | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )  | Remarks  | Location<br>Code |
|--|--|---|--|------------------|
| <b>TORQUE</b>  |  |   | Note 1: Values may be given in other units, i.e. lbf/ft, mV/V etc. | Lab 2            |
| Torque Wrenches and screwdrivers   | BS EN 6789-2:2017 (See note 2)<br>0.1 to 1040 N·m  | 1.0 %   | Note 2: Excluding setting screwdrivers below 1 N·m at lowest level |                  |
|  | BS EN 6789:2003 (withdrawn)<br>0.1 to 2500 N·m   | 1.6 %   |  |                  |
| Static Torque Transducers  | 0.005 to 1500 N·m<br>Classes 0.05 to 5.0<br>See Note 1   | 0.040 % of reading  | Calibrated to BS 7882:2017   |                  |
| <b>TEMPERATURE</b>   |  |   |  | Lab 2            |
| Temperature block calibrators  | -50 °C to +100 °C<br>100 °C to 250 °C<br>250 °C to 650 °C  | 0.060 °C<br>0.060 °C to 0.13 °C<br>0.13 °C  |  |                  |
| Resistance thermometers  | -40 °C to 0 °C<br>0 °C<br>Ambient temperature (20 °C)<br>0 °C to 250 °C<br>250 °C to 270 °C<br>270 °C to 650 °C  | 0.040 °C<br>0.025 °C<br>0.025 °C<br>0.040 °C<br>0.090 °C<br>0.35 °C   | Comparison in liquid bath, ice bath or block calibrator            |                  |
| Thermocouples  | -40 °C to +50 °C<br>50 °C to 250 °C<br>250 °C to 650 °C<br>650 °C to 1000 °C<br>1000 °C to 1200 °C   | 0.10 °C<br>0.10 °C to 0.20 °C<br>0.35 °C to 0.60 °C<br>1.7 °C to 2.5 °C<br>2.5 °C to 3.5 °C                 | Comparison in liquid bath, ice bath or block calibrator            |                  |
| Thermocouples, batch calibration   | -40 °C to +250 °C  | 0.080 °C  | Comparison in liquid bath or ice bath                              |                  |
| Compensating and extension cables for<br>Base metal thermocouples<br>Noble metal thermocouples | 0 °C and 20 °C to 40 °C<br>0 °C and 20 °C to 40 °C   | 0.055 °C<br>0.12 °C   | Comparison in liquid bath or ice bath                              |                  |
| Digital thermometers with probes   | -40 °C to 0 °C<br>0 °C<br>Ambient temperature (20 °C)<br>0 °C to 250 °C<br>250 °C to 270 °C<br>270 °C to 650 °C<br>650 °C to 1000 °C<br>1000 °C to 1200 °C | 0.040 °C<br>0.025 °C<br>0.025 °C<br>0.040 °C<br>0.065 °C<br>0.25 °C<br>1.7 °C to 2.5 °C<br>2.5 °C to 3.5 °C | Comparison in liquid bath, ice bath or block calibrator            |                  |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

**Calibration performed by the Organisation at the locations specified**

| Measured Quantity<br>Instrument or Gauge  | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )             | Remarks   | Location<br>Code |
|---|---|--|---|------------------|
| TEMPERATURE (continued)   |   |  |   |                  |
| Liquid in glass thermometers  | -40 °C to +70 °C<br>70 °C to 270 °C   | 0.050 °C<br>0.050 °C to 0.20 °C                                | Comparison in liquid bath, ice bath or block calibrator.<br><br>Thermometers can be read with an uncertainty of 1/5 of the smallest marked division |                  |
| Liquid baths  | -50 °C to +300 °C   | 0.040 °C   | Using PRTs. Suitable for characterisation of temperature calibration baths  |                  |
| Temperature controlled environmental chambers, fridges/refrigerators, freezers, ovens, incubators and rooms | -40 °C to +250 °C   | 1.0 °C   | Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping.                           | Site             |
| Digital thermometers with probes  | -20 °C to +140 °C   | 0.20 °C  |   |                  |
| Digital thermometers with probes and indicators including data loggers                                      | 10 °C to 25 °C<br>25 °C to 37 °C<br>37 °C to 50 °C                                  | 0.18 °C to 0.14 °C<br>0.14 °C to 0.15 °C<br>0.15 °C to 0.27 °C | Performed in air  | Lab 2            |
| RELATIVE HUMIDITY   |   |  |   |                  |
| Relative humidity instruments with displays including data loggers  | Performed at 25 °C<br>20 %rh to 80 %rh  | 1.70 %rh   | By comparison with a reference hygrometer   | Lab 2            |
| ELECTRICAL SIMULATION   |   |  |   |                  |
| Electrical calibration of temperature indicators and simulators   |   |  | Uncertainties based on Types J, K, R and B. Uncertainties may be higher for other thermocouple types  | Lab 2            |
| Base metal thermocouples  | -200 °C to -100 °C<br>-100 °C to -30 °C<br>-30 °C to +1200 °C<br>1200 °C to 1370 °C | 0.18 °C<br>0.090 °C<br>0.080 °C<br>0.11 °C                     | Including cold junction compensation  |                  |
| Noble metal thermocouples   | -50 °C to +50 °C<br>50 °C to 270 °C<br>270 °C to 1820 °C                            | 0.70 °C<br>0.50 °C<br>0.40 °C                                  | Including cold junction compensation  |                  |
| Cold junction compensation  | Ambient temperature   | 0.080 °C   |   |                  |





0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge   | Range  | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )   | Remarks                                     | Location<br>Code                      |
|--|--|--|---|---------------------------------------|
| <p>PRESSURE</p> <p>Gas pressure (absolute)</p> <p>Calibration of pressure measuring instruments and gauges</p> <p>Gas pressure (gauge)</p> <p>Calibration of pressure measuring instruments and gauges</p> <p>Hydraulic pressure (gauge)</p> <p>Calibration of pressure measuring instruments and gauges</p> <p>Hydraulic pressure (gauge)</p> <p>Calibration of pressure measuring instruments and gauges</p> | <p>3.5 kPa to 175 kPa<br/>175 kPa to 700 kPa<br/>700 kPa to 7 MPa<br/>7 MPa to 17 MPa</p> <p>-100 kPa to 0 kPa<br/>0 kPa to 3.5 kPa<br/>3.5 kPa to 175 kPa<br/>175 kPa to 700 kPa<br/>700 kPa to 7 MPa<br/>7 MPa to 17 MPa</p> <p>350 kPa to 0.69 MPa<br/>0.70 MPa to 7.1 MPa<br/>7.1 MPa to 83.5 MPa<br/>83.5 MPa to 140 MPa</p> <p>0 MPa to 7 MPa<br/>7 MPa to 35 MPa<br/>35 MPa to 70 MPa<br/>70 MPa to 140 MPa</p> | <p>0.0094 % + 2.0 Pa<br/>0.010 % + 0.020 Pa<br/>0.010 % + 0.20 Pa<br/>0.028 % + 73 Pa</p> <p>97 Pa<br/>2.0 Pa<br/>0.0090 % + 0.020 Pa<br/>0.0096 % + 0.020 Pa<br/>0.0096 % + 0.20 Pa<br/>0.028 %</p> <p>0.0082 % + 0.13 kPa<br/>0.0094 % + 0.11 kPa<br/>0.0090 % + 0.11 kPa<br/>0.080 % + 10 kPa</p> <p>0.080 % + 0.10 kPa<br/>0.090 % + 1.0 kPa<br/>0.20 % + 7.0 kPa<br/>0.080 % + 10 kPa</p> | <p>Methods consistent with EURAMET CG17</p> | <p>Lab 2</p> <p>Lab 2</p> <p>Site</p> |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge  | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ ) | Remarks   | Location<br>Code |
|---|---|--|---|------------------|
| NON AUTOMATIC<br>WEIGHING INSTRUMENTS<br>(NAWI)   | Capacity  |  | Note 1: Weights are available<br>in   | Site             |
| Electronic and mechanical<br>self-indicating non automatic<br>weighing instruments and<br>spring balances<br>See notes 1 and 2  | 200 mg  | 0.012 mg   | OIML Class E2<br>Max grouped load 800 g                                     |                  |
|   | 500 mg  | 0.015 mg   |   |                  |
|   | 1 g   | 0.018 mg   |   |                  |
|   | 2 g   | 0.025 mg   |   |                  |
|   | 5 g   | 0.037 mg   |   |                  |
|   | 10 g  | 0.042 mg   |   |                  |
|   | 12 g  | 0.058 mg   |   |                  |
|   | 50 g  | 0.092 mg   |   |                  |
|   | 100 g   | 0.16 mg  |   |                  |
|   | 200 g   | 0.30 mg  |   |                  |
|   | 500 g   | 0.78 mg  |   |                  |
|   | 1 kg  | 3.3 mg   | OIML Class F1<br>Max grouped load 95 kg                                     |                  |
|   | 2 kg  | 6.5 mg   |   |                  |
|   | 5 kg  | 16 mg  |   |                  |
|   | 10 kg   | 33 mg  |   |                  |
|   | 20 kg   | 65 mg  |   |                  |
|   | 50 kg   | 160 mg   |   |                  |
|   | 100 kg  | 3.5 g  |   |                  |
|   | 200 kg  | 7.0 g  |   |                  |
|   | 500 kg  | 18   |   |                  |
| 1000 kg   | 35 g  |  |   |                  |
| 2000 kg   | 70 g  | OIML Class M1<br>Max grouped load 3500 kg          |   |                  |
| 3500 kg   | 110 g   |  |   |                  |
|   |   |  | Note 2: Other loads within the<br>overall listed range may also<br>be used. |                  |
|   |   |  | Note 3: The calibration<br>procedure PRO17 is based<br>on EURAMET CG-18     |                  |
| FORCE   |   |  |   | Lab 2            |
| Calibration of force measuring<br>devices (e.g. strain gauged<br>load cells and load measuring<br>rings) but excluding proving<br>devices, in tension and<br>compression modes using<br>masses and reference<br>devices | From 0.1 N up to 2500 N<br>From 0.6 kN up to 600 kN | 0.10 %<br>0.23 %                                   | Calibration performed to BS<br>8422:2003                                    |                  |
| Calibration of push pull force<br>measuring devices in tension<br>and compression using<br>masses and reference<br>devices  | 0.001 N to 2 kN                                     | 0.10 %   |   |                  |
| END   |   |  |   |                  |



0666  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Allied Aerosystems Ltd**  
**Issue No: 058 Issue date: 03 September 2024**

Calibration performed by the Organisation at the locations specified

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$